REVIEW

Recent advances in echocardiography for valvular heart disease [version 1; peer review: 3 approved]

Rebecca Hahn

Columbia University Medical Center/New York Presbyterian Hospital, New York, NY, USA

V1 First published: 28 Sep 2015, 4(F1000 Faculty Rev):914
<https://doi.org/10.12688/f1000research.6446.1>
Latest published: 28 Sep 2015, 4(F1000 Faculty Rev):914
<https://doi.org/10.12688/f1000research.6446.1>

Abstract

Echocardiography is the imaging modality of choice for the assessment of patients with valvular heart disease. Echocardiographic advancements may have particular impact on the assessment and management of patients with valvular heart disease. This review will summarize the current literature on advancements, such as three-dimensional echocardiography, strain imaging, intracardiac echocardiography, and fusion imaging, in this patient population.

Keywords

echocardiography, valvular heart disease, three-dimensional echocardiography, strain imaging, intracardiac echocardiography, fusion imaging

Open Peer Review

Approval Status

	1	2	3
version 1			
28 Sep 2015			
1. Judy Hung , Massachusetts General Hospital, Boston, USA			
2. Jeroen J. Bax , Leiden University Medical Center, Leiden, The Netherlands			
3. Takahiro Shiota , Cedars-Sinai Heart Institute, Los Angeles, USA			

Any comments on the article can be found at the end of the article.

Corresponding author: Rebecca Hahn (rth2@cumc.columbia.edu)

Competing interests: Rebecca T. Hahn has received speaker honoraria from Edwards Lifesciences, St. Jude Medical and Boston Scientific, and a research grant from Philips Healthcare.

Grant information: The author(s) declared that no grants were involved in supporting this work.

Copyright: © 2015 Hahn R. This is an open access article distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

How to cite this article: Hahn R. **Recent advances in echocardiography for valvular heart disease [version 1; peer review: 3 approved]** F1000Research 2015, 4(F1000 Faculty Rev):914 <https://doi.org/10.12688/f1000research.6446.1>

First published: 28 Sep 2015, 4(F1000 Faculty Rev):914 <https://doi.org/10.12688/f1000research.6446.1>

Introduction

The American Heart Association/American College of Cardiology Guidelines for the Management of Patients with Valvular Heart Disease¹ state that echocardiography (transthoracic [TTE] or transesophageal [TEE]) is the imaging modality of choice for the assessment of patients with valvular heart disease. Numerous less invasive therapies, such as percutaneous or transcatheter interventions, have recently been introduced for the treatment of structural heart disease. Many of these procedures require extensive multi-modality imaging guidance and have increased interest in advancements in echocardiography. Recent advancements in echocardiography are of particular relevance to valvular disease. This review will discuss the application of these new technologies to diagnose and manage various types of valvular heart diseases.

Three-dimensional echocardiography

The echocardiographic advancement that has had the most impact on the diagnosis of valvular heart disease is real time three-dimensional (RT3D) echocardiography. The advantages of three-dimensional (3D) imaging over two-dimensional (2D) imaging has been well described in the most recent societal guidelines: “Recommendations for cardiac chamber quantification by echocardiography in adults” update² and “Recommendations for image acquisition and display using 3D echocardiography”³. These guidelines review the significant data supporting the improved accuracy and reproducibility of 3D imaging for ventricular volumes and mass, as well as valvular morphology and function. Initially introduced in the year 2000, the continued improvement of 3D technology has led to its widespread availability and its growing utility, particularly for valvular heart disease⁴.

Valve morphology

RT3D TEE has significantly changed the assessment of valvular pathology and has revolutionized patient selection not only for surgical repair but also for newer transcatheter procedures, discussed in a subsequent section.

RT3D TEE is not only more accurate than 2D techniques in identifying specific mitral valve pathology in the setting of complex disease but the diagnosis can be made more rapidly, which is of particular use in the intraoperative evaluation of patients undergoing mitral valve repair^{5–9}. RT3D echo improves the accuracy and reproducibility of planimetry measurements of mitral valve area in the setting of rheumatic disease by ensuring on-axis imaging of the short-axis view^{10–13}. This technology has also been integral to our understanding of the dynamic nature of the mitral valve complex in normal patients, as well as in primary and secondary mitral valve disease^{14–18}.

Recent RT3D TEE studies have shown a coupling of mitral and aortic valve dynamic anatomy. Mitral valve diseases may affect normal mitral-aortic coupling and aortic valve function; different patterns of abnormal mitral-aortic coupling are associated with different Carpentier types of mitral regurgitation¹⁹. Conversely, changes in aortic morphology may affect mitral valve function, particularly in the setting of aortic stenosis and calcification of the aortic-mitral fibrous continuity²⁰. RT3D TEE has shown changes in mitral valve morphology following surgical aortic valve replacement²¹ as well as

transcatheter aortic valve replacement (TAVR)²². Notably, a decrease of tenting area predicted those patients whose mitral regurgitation improved following TAVR.

Aortic valve morphology as well as aortic root measurements are more accurate and reproducible with RT3D imaging. Inter-commissural distance and free leaflet edge lengths can be measured by 3D echocardiography and are used to choose the tube graft size in valve-sparing root operations²³. Larger left ventricular outflow tract areas and calculated aortic valve dimensions and areas are obtained by RT3D TEE²⁴. Planimetry of the aortic valve and left ventricular outflow tract area by RT3D has been shown to be accurate and reproducible^{25–27} and may influence surgical decision-making in the setting of moderate-to-severe aortic stenosis²⁸. With accurate measurement of the left ventricular outflow tract, geometric assumptions used in the continuity equation are avoided, resulting in more precise estimations of aortic valve areas using 3D echocardiography over traditional 2D methods.

Prosthetic valve function can also be accurately assessed using RT3D TEE. With transcatheter solutions to bioprosthetic valve failure^{29–31} and paravalvular regurgitation^{32–34} RT3D TEE has become an important tool for intra-procedural guidance during percutaneous interventions. TEE can depict not only the relevant cardiac landmarks adjacent to the sites of paravalvular leaks but also wires, delivery catheters, and closure devices³⁵. RT3D TEE imaging results in a more accurate localization of paravalvular defects and an estimation of the size of the defect that correlated better with surgical findings when compared with 2D TEE³⁶.

Quantification of valvular function

Three-dimensional color Doppler may overcome the limitations of 2D and standard Doppler measurements for quantifying regurgitation^{3,37,38}. Studies have shown the feasibility of measuring the 3D vena contracta (narrowest portion of the regurgitant jet) on RT3D echocardiography to assess the severity of regurgitation for native regurgitant valve disease^{37,39–41}, as well as following surgical⁴² or transcatheter interventions⁴³.

Calculation of regurgitant volume in native valvular disease using the proximal isovelocity surface area (PISA) method⁴⁴ has known technical limitations, primarily the geometric assumptions of PISA shape required to calculate effective regurgitant orifice area. Multiple studies have validated the use of single-beat RT3D echocardiographic color Doppler imaging allowing the direct measurement of PISA without geometric assumptions for aortic, mitral, and tricuspid regurgitation assessment^{45–48}.

Newer methods of determining relative flows within the heart make use of the velocity and direction of flow information which can be derived from color Doppler. Off-line software has been developed which uses 2D color Doppler images to determine the velocity, flow rate, and flow volume in any given region of the heart⁴⁹. Extension of this technology to 3D color Doppler volume sets is now possible and allows rapid, accurate, and reproducible quantitation of relative stroke volumes^{50,51}. Thavendiranathan *et al.*⁵¹ used the velocity information encoded in the volume color Doppler data, targeting the appropriate region of interest by using the simultaneous 3D imaging of the

mitral annulus and left ventricular outflow tract. Color Doppler velocity is multiplied by a known area of this cross-section (a voxel area), and the resulting spatially averaged flow rates are used to generate flow-time curves that resemble those obtained by magnetic resonance imaging. The temporal integration of the flow-time curve generates the stroke volume. There was excellent correlation between the automated measured mitral inflow and aortic stroke volumes, and magnetic resonance imaging stroke volume ($r = 0.91$, 95% confidence interval [CI], 0.83–0.95, and $r = 0.93$, 95% CI, 0.87–0.96, respectively, $P < 0.001$) and very low interobserver variability. Automation of the measurement process allowed calculations of mitral inflow and aortic stroke volumes to be performed very rapidly. This methodology will likely become the standard for measurement of regurgitant volumes in the future.

Structural heart disease interventions

TAVR has become an acceptable alternative treatment for high-risk or inoperable patients with severe symptomatic aortic stenosis^{52–55}. Three-dimensional echocardiography has been shown to improve sizing of the transcatheter valve^{56–58}. RT3D TEE is comparable to computed tomography for annular assessment and prediction of paravalvular regurgitation due to oversizing^{59,60}, as well as measurement of coronary artery height⁶¹. RT3D TEE has been shown

to provide superior spatial visualization and anatomic orientation, optimizing procedural performance, and RT3D TTE can be used to assess the severity of paravalvular regurgitation following TAVR⁶². Further study of this technique for quantifying regurgitant severity is warranted in addition to a unified scheme for grading paravalvular regurgitation following TAVR⁶³. Newer devices, with features such as external skirts or the ability to reposition, may reduce the incidence of post-TAVR complication.

Three-dimensional TEE may also improve procedural success and shorten procedure time for the MitraClip™ device (Abbott Vascular Structural Heart, Menlo Park, CA) (Figure 1)^{64–66}. Altiok *et al.*⁶⁵ performed a structured analysis to compare information and guidance capability provided by RT3D TEE compared to 2D TEE and found 3D TEE advantageous in 9 of 11 steps of the percutaneous mitral repair procedure, including optimizing trans-septal puncture site, guidance of the clip delivery system, precise positioning of the clip delivery system simultaneously in anterior-posterior and lateral-medial direction, valvular regurgitation jet position, adjustment and visualization of the clip position relative to the valvular orifice, and assessment of remaining regurgitant jets⁶⁵. Following MitraClip, assessment of residual regurgitation could also be assessed by 3D color Doppler⁴³. A >50% reduction in regurgitant

Figure 1. Three-dimensional echocardiography during a transcatheter mitral repair procedure. Panel A shows the baseline mitral valve morphology with a very large prolapsing and partially flail P2 (middle) scallop (red arrows). Panel B shows positioning of the MitraClip device (blue arrow). Panel C is a dual plane three-dimensional image (ventricular and atrial views) of the final, 2-clip (yellow arrows) resulting double orifice. There was trivial residual mitral regurgitation.

volume using the product of vena contracta areas defined by direct planimetry of RT3D color Doppler and velocity time integral using continuous-wave Doppler was associated with greater left atrial and ventricular remodeling.

Strain imaging

Recent American Society of Echocardiography Chamber Quantification guidelines strongly recommend routine assessment of ventricular systolic function by quantification of ventricular volumes and calculation of ejection fraction (EF)². Cardiac mechanics, however, can now be assessed with the use of both tissue Doppler and speckle tracking for the measurement of myocardial displacement⁶⁷. The measurement of myocardial deformation or “strain” is the fractional change in the length of a myocardial segment (expressed as a percentage of the baseline length). Strain rate is the rate of change in strain. The deformation of the myocardium is directional: lengthening would be represented by positive strain, and shortening by negative strain. Systolic strain can be measured along the anatomic coordinates of the cardiac chambers: longitudinal (negative strain), radial (positive strain), and circumferential (negative strain). The strengths and weaknesses of strain measurement have been well described⁶⁷; however, the recent standardization of strain Digital Imaging and Communications in Medicine (DICOM) format will reduce inter-vendor variability which, along with improved software analysis and automation packages, will likely increase the clinical acceptability and use of this powerful technique.

Aortic valve disease

Numerous studies have shown the utility of strain imaging for assessing left ventricular function in aortic valve disease. In the presence of normal EF, increasing severity of aortic stenosis was associated with reduced global longitudinal strain (GLS)^{68,69}. Subclinical improvement in global and regional systolic function by tissue Doppler and speckle strain also occurs following TAVR (Figure 2)^{70–72}. In low flow, low gradient, severe aortic stenosis with normal EF, strain parameters improved following TAVR, even in the absence of significant change in EF⁷³. Regional strain abnormalities in patients with severe aortic stenosis may be able to further sub-stratify patients

with concomitant infiltrative diseases, such as amyloid as well as coronary disease. In patients with cardiac amyloid, relative apical sparing (with preserved apical longitudinal strain) was sensitive (93%) and specific (82%) in differentiating amyloid from controls, some of whom had severe aortic stenosis. In patients with moderate or severe aortic stenosis and concomitant coronary disease, on the other hand, worse apical and mid longitudinal strain parameters were predictive of significant coronary artery stenosis⁷⁴.

Because mortality is significantly associated with symptom development⁷⁵, strain has been postulated as a possible early marker of ventricular dysfunction in asymptomatic patients with severe aortic stenosis and thus may be a useful tool in determining the timing of intervention in this population. In fact, Carasso *et al.*⁷⁶ showed that longitudinal strain was low in asymptomatic patients with severe aortic stenosis with supernormal apical circumferential strain and rotation. In symptomatic patients, however, longitudinal strain was significantly lower with no compensatory circumferential myocardial mechanics. Other investigators suggest that, after adjusting for aortic stenosis severity and EF, only basal longitudinal strain (and not GLS) was an independent predictor of symptomatic status⁷⁷. In fact, following TAVR, the improvement in GLS may be a result of basal and mid segment improvement only⁷⁸.

Strain imaging may be particularly useful in predicting outcomes in patients with severe aortic stenosis. In patients with low flow, low gradient, aortic stenosis with normal EF, a recent study showed both stroke volume index (≤ 35 ml/m²) and GLS ($> -15\%$) are independently associated with worse survival⁷⁹. In patients with low flow, low gradient, aortic stenosis with reduced EF, GLS is independently associated with mortality and dobutamine stress GLS may provide incremental prognostic value beyond GLS measured at rest⁸⁰. Three-dimensional GLS may be a better predictor of outcome compared to 2D strain⁸¹. Finally, Kusunose *et al.*⁸² studied 395 patients with moderate-severe aortic stenosis (aortic valve area < 1.3 cm²) and found that GLS was an independent predictor of mortality in this population. A GLS $> -12\%$ was associated with the lowest survival⁸².

Figure 2. Strain imaging during transcatheter aortic valve replacement (TAVR). Panel A shows a global circumferential strain (GCS) of -11% prior to TAVR. Panel B shows a GCS of -18% following TAVR. This represents an improvement (greater shortening) in ventricular mechanics.

Deformation characteristics have also been studied in patients with aortic regurgitation⁸³⁻⁸⁷. In a prospective study of young patients (<18 years old) with aortic regurgitation, the only significant predictor of progression of disease on multi-variable analysis was GLS ($P=0.04$, cut-off value of $>-19.5\%$, sensitivity of 77.8% , specificity of 94.1% , and area under the curve of 0.89)⁸³. Prospective studies of adult patients have also shown that strain parameters by speckle-tracking could detect early myocardial systolic and diastolic dysfunction, and lower strain values were associated with disease progression in medically managed patients, or impaired outcomes in surgically treated patients⁸⁵. A systolic radial strain rate of $<1.82/\text{sec}$ was a good predictor of postoperative left ventricular dysfunction⁸⁶. Finally, in a prospective study, 60 patients with chronic aortic regurgitation were followed for 64 months and global longitudinal strain (four-chamber view only) was an independent predictor of mortality (hazard ratio 1.313 , 95% CI $1.010-1.706$, $P=0.042$)⁸⁷.

Mitral valve disease

Chronic mitral regurgitation is associated with complex left ventricular adaptive remodeling, eccentric hypertrophy, and, eventually, reduced EF. Current guidelines recommend intervening on severe, asymptomatic mitral regurgitation in the setting of reduced EF because of a high incidence of persistent or worsening dysfunction⁸⁸. In chronic severe degenerative mitral regurgitation, numerous studies have shown that a reduced baseline GLS signifies a maladaptive preload-related change that is associated with a reduction in left ventricular EF immediately after mitral valve repair⁸⁹⁻⁹¹. A GLS cutoff of $>-19.9\%$ was a strong independent predictor of long-term left ventricular dysfunction and may become an appropriate indication for intervention in the setting of normal EF⁹⁰.

Intra-cardiac echocardiography

Although TEE imaging is well established and provides exceptional images, particularly for intra-procedural guidance, it most

commonly requires general anesthesia and may be associated with intermittent obstruction of fluoroscopic viewing⁹². With the current move toward conscious sedation for structural heart disease interventions, intra-cardiac echocardiography (ICE) may be an acceptable alternative in some patients with no other adequate intra-procedural imaging options. Evidence that ICE guidance can improve safety and outcome of interventional procedures is still lacking; however, ICE imaging for paravalvular leak closure has been reported to be feasible and advantageous^{32,93}. A reduction in contrast use has also been reported with 2D ICE when used in TAVR (Figure 3)⁹⁴. The recently introduced AcuNav® V catheter (Siemens Inc. Mountain View, USA) represents the only commercially available RT3D ICE system. The 10F catheter carries a matrix transducer providing a $22^\circ \times 90^\circ$ real-time volume image. This small volume represents the main limitation, particularly in near field applications, such as structural heart disease interventions.

Fusion imaging

Combining images from two or more different imaging techniques, or fusion imaging, has been accomplished most recently with real-time echocardiography and fluoroscopy⁹⁵⁻⁹⁷. This technology, which co-registers the TEE probe position with the intervention table and the angulation of the fluoroscopy C-arm, allows for relatively accurate placement of the TEE image onto the fluoroscopic image. This integration eliminates the need for two different image display monitors and the mental integration of two very different imaging datasets by the operator of structural heart disease interventions.

The ability to define targets on echocardiographic images (whether 2D or 3D), and co-register these targets onto the fluoroscopic images, should improve guidance of structural heart disease interventions (Figure 4). This technology has been shown to be safe and feasible

Figure 3. Intra-cardiac echocardiography (ICE) during transcatheter aortic valve replacement (TAVR). Panel A is the two-dimensional ICE image with panel B showing the simultaneous three-dimensional volume during positioning of the transcatheter aortic valve (yellow arrows).

Figure 4. Hybrid/fusion imaging during paravalvular regurgitation closure. After coregistration of the transesophageal echocardiographic probe with the fluoroscopic image, the two images can be fused to allow a more comprehensive understanding of anatomy. Localizing the regurgitant orifice on transesophageal echo imaging can then be translated to the corresponding position on the fluoroscopic image.

for the transcatheter mitral repair procedure with the MitraClip™ device (Abbott Vascular Structural Heart, Menlo Park, CA) and shows a trend towards reduction of fluoroscopy and procedure time⁹⁸.

Conclusion

Echocardiography is the primary imaging modality for the diagnosis and management of patients with valvular heart disease. Improvement in surgical outcomes and advances in interventional techniques require further refinements in echocardiographic imaging. Three-dimensional echocardiography, strain imaging, intracardiac echocardiography, and fusion imaging have significant application in advancing our understanding of pathophysiology and anatomy, as well as the diagnosis and management of patients with valvular heart disease.

Abbreviations

2D, two-dimensional; 3D, three-dimensional; confidence interval, CI; EF, ejection fraction; GLS, global longitudinal strain; ICE, intra-cardiac echocardiography; PISA, proximal isovelocity surface area; RT3D, real time three-dimensional; TAVR, transcatheter aortic valve replacement; TEE, transesophageal echocardiography; TTE, transthoracic echocardiography.

Competing interests

Rebecca T. Hahn has received speaker honoraria from Edwards Lifesciences, St. Jude Medical and Boston Scientific, and a research grant from Philips Healthcare.

Grant information

The author(s) declared that no grants were involved in supporting this work.

References

1. Nishimura RA, Otto C: **2014 ACC/AHA valve guidelines: earlier intervention for chronic mitral regurgitation.** *Heart.* 2014; **100**(12): 905–7.
[PubMed Abstract](#) | [Publisher Full Text](#)
2. Lang RM, Badano LP, Mor-Avi V, *et al.*: **Recommendations for cardiac chamber quantification by echocardiography in adults: an update from the American Society of Echocardiography and the European Association of Cardiovascular Imaging.** *J Am Soc Echocardiogr.* 2015; **28**(1): 1–39.e14.
[PubMed Abstract](#) | [Publisher Full Text](#)
3. Lang RM, Badano LP, Tsang W, *et al.*: **EAE/ASE recommendations for image acquisition and display using three-dimensional echocardiography.** *J Am Soc Echocardiogr.* 2012; **25**(1): 3–46.
[PubMed Abstract](#) | [Publisher Full Text](#)
4. Lang RM, Badano LP, Tsang W, *et al.*: **EAE/ASE recommendations for image acquisition and display using three-dimensional echocardiography.** *Eur Heart J Cardiovasc Imaging.* 2012; **13**(1): 1–46.
[PubMed Abstract](#) | [Publisher Full Text](#)
5. Grewal J, Mankad S, Freeman WK, *et al.*: **Real-time three-dimensional transesophageal echocardiography in the intraoperative assessment of mitral valve disease.** *J Am Soc Echocardiogr.* 2009; **22**(1): 34–41.
[PubMed Abstract](#) | [Publisher Full Text](#)
6. Pepi M, Tamborini G, Maltagliati A, *et al.*: **Head-to-head comparison of two- and three-dimensional transthoracic and transesophageal echocardiography in the localization of mitral valve prolapse.** *J Am Coll Cardiol.* 2006; **48**(12): 2524–30.
[PubMed Abstract](#) | [Publisher Full Text](#)
7. Ben Zekry S, Nagueh SF, Little SH, *et al.*: **Comparative accuracy of two- and three-dimensional transthoracic and transesophageal echocardiography in identifying mitral valve pathology in patients undergoing mitral valve repair: initial observations.** *J Am Soc Echocardiogr.* 2011; **24**(10): 1079–85.
[PubMed Abstract](#) | [Publisher Full Text](#)
8. Hien MD, Rauch H, Lichtenberg A, *et al.*: **Real-time three-dimensional transesophageal echocardiography: improvements in intraoperative mitral valve imaging.** *Anesth Analg.* 2013; **116**(2): 287–95.
[PubMed Abstract](#) | [Publisher Full Text](#)
9. Levack MM, Jassar AS, Shang EK, *et al.*: **Three-dimensional echocardiographic analysis of mitral annular dynamics: implication for annuloplasty selection.** *Circulation.* 2012; **126**(11 Suppl 1): S183–8.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
10. Zamorano J, Cordeiro P, Sugeng L, *et al.*: **Real-time three-dimensional echocardiography for rheumatic mitral valve stenosis evaluation: an accurate and novel approach.** *J Am Coll Cardiol.* 2004; **43**(11): 2091–6.
[PubMed Abstract](#) | [Publisher Full Text](#)
11. Schlosshan D, Aggarwal G, Mathur G, *et al.*: **Real-time 3D transesophageal echocardiography for the evaluation of rheumatic mitral stenosis.** *JACC Cardiovasc Imaging.* 2011; **4**(6): 580–8.
[PubMed Abstract](#) | [Publisher Full Text](#)
12. Min SY, Song JM, Kim YJ, *et al.*: **Discrepancy between mitral valve areas measured by two-dimensional planimetry and three-dimensional transesophageal echocardiography in patients with mitral stenosis.** *Heart.* 2013; **99**(4): 253–8.
[PubMed Abstract](#) | [Publisher Full Text](#)
13. Wunderlich NC, Beigel R, Siegel RJ: **Management of mitral stenosis using 2D and 3D echo-Doppler imaging.** *JACC Cardiovasc Imaging.* 2013; **6**(11): 1191–205.
[PubMed Abstract](#) | [Publisher Full Text](#)
14. Levine RA, Handschumacher MD, Sanfilippo AJ, *et al.*: **Three-dimensional echocardiographic reconstruction of the mitral valve, with implications for the diagnosis of mitral valve prolapse.** *Circulation.* 1989; **80**(3): 589–98.
[PubMed Abstract](#) | [Publisher Full Text](#)
15. Topilsky Y, Vaturi O, Watanabe N, *et al.*: **Real-time 3-dimensional dynamics of functional mitral regurgitation: a prospective quantitative and mechanistic study.** *J Am Heart Assoc.* 2013; **2**(3): e000039.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
16. Grewal J, Suri R, Mankad S, *et al.*: **Mitral annular dynamics in myxomatous valve disease: new insights with real-time 3-dimensional echocardiography.** *Circulation.* 2010; **121**(12): 1423–31.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
17. Watanabe N, Ogasawara Y, Yamaura Y, *et al.*: **Mitral annulus flattens in ischemic mitral regurgitation: geometric differences between inferior and anterior myocardial infarction: a real-time 3-dimensional echocardiographic study.** *Circulation.* 2005; **112**(9 Suppl): I458–62.
[PubMed Abstract](#)
18. Zeng X, Nunes MCP, Dent J, *et al.*: **Asymmetric versus symmetric tethering patterns in ischemic mitral regurgitation: geometric differences from three-dimensional transesophageal echocardiography.** *J Am Soc Echocardiogr.* 2014; **27**(4): 367–75.
[PubMed Abstract](#) | [Publisher Full Text](#)
19. Looi JL, Lee AP, Fang F, *et al.*: **Abnormal mitral-aortic intervalvular coupling in mitral valve diseases: a study using real-time three-dimensional transesophageal echocardiography.** *Clin Res Cardiol.* 2015.
[PubMed Abstract](#) | [Publisher Full Text](#)
20. Tsang W, Meineri M, Hahn RT, *et al.*: **A three-dimensional echocardiographic study on aortic-mitral coupling in transcatheter aortic valve replacement.** *Eur Heart J Cardiovasc Imaging.* 2013; **14**(10): 950–6.
[PubMed Abstract](#) | [Publisher Full Text](#)
21. Mahmood F, Warraich HJ, Gorman JH 3rd, *et al.*: **Changes in mitral annular geometry after aortic valve replacement: a three-dimensional transesophageal echocardiographic study.** *J Heart Valve Dis.* 2012; **21**(6): 696–701.
[PubMed Abstract](#) | [Free Full Text](#)
22. Shibayama K, Harada K, Berdejo J, *et al.*: **Effect of transcatheter aortic valve replacement on the mitral valve apparatus and mitral regurgitation: real-time three-dimensional transesophageal echocardiography study.** *Circ Cardiovasc imaging.* 2014; **7**(2): 344–51.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
23. Otani K, Takeuchi M, Kaku K, *et al.*: **Assessment of the aortic root using real-time 3D transesophageal echocardiography.** *Circ J.* 2010; **74**(12): 2649–57.
[PubMed Abstract](#) | [Publisher Full Text](#)
24. Saitoh T, Shiota M, Izumo M, *et al.*: **Comparison of left ventricular outflow geometry and aortic valve area in patients with aortic stenosis by 2-dimensional versus 3-dimensional echocardiography.** *Am J Cardiol.* 2012; **109**(11): 1626–31.
[PubMed Abstract](#) | [Publisher Full Text](#)
25. Goland S, Trento A, Iida K, *et al.*: **Assessment of aortic stenosis by three-dimensional echocardiography: an accurate and novel approach.** *Heart.* 2007; **93**(7): 801–7.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
26. Doddamani S, Bello R, Friedman MA, *et al.*: **Demonstration of left ventricular outflow tract eccentricity by real time 3D echocardiography: implications for the determination of aortic valve area.** *Echocardiography.* 2007; **24**(8): 860–6.
[PubMed Abstract](#) | [Publisher Full Text](#)
27. Shahgaldi K, Manouras A, Brodin LÅ, *et al.*: **Direct measurement of left ventricular outflow tract area using three-dimensional echocardiography in biplane mode improves accuracy of stroke volume assessment.** *Echocardiography.* 2010; **27**(9): 1078–85.
[PubMed Abstract](#) | [Publisher Full Text](#)
28. Jainandunsing JS, Mahmood F, Matyal R, *et al.*: **Impact of three-dimensional echocardiography on classification of the severity of aortic stenosis.** *Ann Thorac Surg.* 2013; **96**(4): 1343–8.
[PubMed Abstract](#) | [Publisher Full Text](#)
29. Dvir D, Barbanti M, Tan J, *et al.*: **Transcatheter aortic valve-in-valve implantation for patients with degenerative surgical bioprosthetic valves.** *Curr Probl Cardiol.* 2014; **39**(1): 7–27.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
30. Cheung A, Webb JG, Barbanti M, *et al.*: **5-year experience with transcatheter transapical mitral valve-in-valve implantation for bioprosthetic valve dysfunction.** *J Am Coll Cardiol.* 2013; **61**(17): 1759–66.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
31. Cullen MW, Cabalka AK, Alli OO, *et al.*: **Transvenous, antegrade Melody valve-in-valve implantation for bioprosthetic mitral and tricuspid valve dysfunction: a case series in children and adults.** *JACC Cardiovasc Interv.* 2013; **6**(6): 598–605.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
32. Rihal CS, Sorajja P, Booker JD, *et al.*: **Principles of percutaneous paravalvular leak closure.** *JACC Cardiovasc Interv.* 2012; **5**(2): 121–30.
[PubMed Abstract](#) | [Publisher Full Text](#)
33. Ruiz CE, Jelnin V, Kronzon I, *et al.*: **Clinical outcomes in patients undergoing percutaneous closure of periprosthetic paravalvular leaks.** *J Am Coll Cardiol.* 2011; **58**(21): 2210–7.
[PubMed Abstract](#) | [Publisher Full Text](#)
34. Nietlispach F, Johnson M, Moss RR, *et al.*: **Transcatheter closure of paravalvular defects using a purpose-specific occluder.** *JACC Cardiovasc Interv.* 2010; **3**(7): 759–65.
[PubMed Abstract](#) | [Publisher Full Text](#)
35. Marsan NA, Tops LF, Nihoyannopoulos P, *et al.*: **Real-time three dimensional echocardiography: current and future clinical applications.** *Heart.* 2009; **95**(22): 1881–90.
[PubMed Abstract](#) | [Publisher Full Text](#)
36. Singh P, Manda J, Hsiung MC, *et al.*: **Live/real time three-dimensional transesophageal echocardiographic evaluation of mitral and aortic valve prosthetic paravalvular regurgitation.** *Echocardiography.* 2009; **26**(8): 980–7.
[PubMed Abstract](#) | [Publisher Full Text](#)
37. Perez de Isla L, Zamorano J, Fernandez-Golfín C, *et al.*: **3D color-Doppler echocardiography and chronic aortic regurgitation: a novel approach for severity assessment.** *Int J Cardiol.* 2013; **166**(3): 640–5.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
38. Gonçalves A, Almeria C, Marcos-Alberca P, *et al.*: **Three-dimensional echocardiography in paravalvular aortic regurgitation assessment after transcatheter aortic valve implantation.** *J Am Soc Echocardiogr.* 2012; **25**(1): 47–55.
[PubMed Abstract](#) | [Publisher Full Text](#)

39. Mori Y, Shiota T, Jones M, *et al.*: **Three-dimensional reconstruction of the color Doppler-imaged vena contracta for quantifying aortic regurgitation: studies in a chronic animal model.** *Circulation.* 1999; 99(12): 1611–7.
[PubMed Abstract](#) | [Publisher Full Text](#)
40. Fang L, Hsiung MC, Miller AP, *et al.*: **Assessment of aortic regurgitation by live three-dimensional transthoracic echocardiographic measurements of vena contracta area: usefulness and validation.** *Echocardiography.* 2005; 22(9): 775–81.
[PubMed Abstract](#) | [Publisher Full Text](#)
41. **F** Chen TE, Kwon SH, Enriquez-Sarano M, *et al.*: **Three-dimensional color Doppler echocardiographic quantification of tricuspid regurgitation orifice area: comparison with conventional two-dimensional measures.** *J Am Soc Echocardiogr.* 2013; 26(10): 1143–52.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
42. **F** Franco E, Almería C, de Agustín JA, *et al.*: **Three-dimensional color Doppler transesophageal echocardiography for mitral paravalvular leak quantification and evaluation of percutaneous closure success.** *J Am Soc Echocardiogr.* 2014; 27(11): 1153–63.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
43. Altioek E, Hamada S, Brehmer K, *et al.*: **Analysis of procedural effects of percutaneous edge-to-edge mitral valve repair by 2D and 3D echocardiography.** *Circ Cardiovasc Imaging.* 2012; 5(6): 748–55.
[PubMed Abstract](#) | [Publisher Full Text](#)
44. Zoghbi WA, Enriquez-Sarano M, Foster E, *et al.*: **Recommendations for evaluation of the severity of native valvular regurgitation with two-dimensional and Doppler echocardiography.** *J Am Soc Echocardiogr.* 2003; 16(7): 777–802.
[PubMed Abstract](#) | [Publisher Full Text](#)
45. **F** Pirat B, Little SH, Igo SR, *et al.*: **Direct measurement of proximal isovelocity surface area by real-time three-dimensional color Doppler for quantitation of aortic regurgitant volume: an *in vitro* validation.** *J Am Soc Echocardiogr.* 2009; 22(3): 306–13.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | [F1000 Recommendation](#)
46. **F** Little SH, Igo SR, Pirat B, *et al.*: ***In vitro* validation of real-time three-dimensional color Doppler echocardiography for direct measurement of proximal isovelocity surface area in mitral regurgitation.** *Am J Cardiol.* 2007; 99(10): 1440–7.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | [F1000 Recommendation](#)
47. **F** de Agustín JA, Marcos-Alberca P, Fernandez-Goffin C, *et al.*: **Direct measurement of proximal isovelocity surface area by single-beat three-dimensional color Doppler echocardiography in mitral regurgitation: a validation study.** *J Am Soc Echocardiogr.* 2012; 25(8): 815–23.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
48. **F** de Agustín JA, Viliani D, Vieira C, *et al.*: **Proximal isovelocity surface area by single-beat three-dimensional color Doppler echocardiography applied for tricuspid regurgitation quantification.** *J Am Soc Echocardiogr.* 2013; 26(9): 1063–72.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
49. Li C, Zhang J, Li X, *et al.*: **Quantification of chronic aortic regurgitation by vector flow mapping: a novel echocardiographic method.** *Eur J Echocardiogr.* 2010; 11(2): 119–24.
[PubMed Abstract](#) | [Publisher Full Text](#)
50. **F** Little SH, Igo SR, McCulloch M, *et al.*: **Three-dimensional ultrasound imaging model of mitral valve regurgitation: design and evaluation.** *Ultrasound Med Biol.* 2008; 34(4): 647–54.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | [F1000 Recommendation](#)
51. **F** Thavendiranathan P, Liu S, Datta S, *et al.*: **Automated quantification of mitral inflow and aortic outflow stroke volumes by three-dimensional real-time volume color-flow Doppler transthoracic echocardiography: comparison with pulsed-wave Doppler and cardiac magnetic resonance imaging.** *J Am Soc Echocardiogr.* 2012; 25(1): 56–65.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
52. **F** Leon MB, Smith CR, Mack M, *et al.*: **Transcatheter aortic-valve implantation for aortic stenosis in patients who cannot undergo surgery.** *N Engl J Med.* 2010; 363(17): 1597–607.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
53. **F** Smith CR, Leon MB, Mack MJ, *et al.*: **Transcatheter versus surgical aortic-valve replacement in high-risk patients.** *N Engl J Med.* 2011; 364(23): 2187–98.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
54. Popma JJ, Adams DH, Reardon MJ, *et al.*: **Transcatheter aortic valve replacement using a self-expanding bioprosthesis in patients with severe aortic stenosis at extreme risk for surgery.** *J Am Coll Cardiol.* 2014; 63(19): 1972–81.
[PubMed Abstract](#) | [Publisher Full Text](#)
55. Adams DH, Popma JJ, Reardon MJ: **Transcatheter aortic-valve replacement with a self-expanding prosthesis.** *N Engl J Med.* 2014; 371(10): 967–8.
[PubMed Abstract](#) | [Publisher Full Text](#)
56. **F** Altioek E, Koos R, Schröder J, *et al.*: **Comparison of two-dimensional and three-dimensional imaging techniques for measurement of aortic annulus diameters before transcatheter aortic valve implantation.** *Heart.* 2011; 97(19): 1578–84.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
57. **F** Gripari P, Ewe SH, Fusini L, *et al.*: **Intraoperative 2D and 3D transesophageal echocardiographic predictors of aortic regurgitation after transcatheter aortic valve implantation.** *Heart.* 2012; 98(16): 1229–36.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
58. Hahn RT, Khaliq O, Williams MR, *et al.*: **Predicting paravalvular regurgitation following transcatheter valve replacement: utility of a novel method for three-dimensional echocardiographic measurements of the aortic annulus.** *J Am Soc Echocardiogr.* 2013; 26(9): 1043–52.
[PubMed Abstract](#) | [Publisher Full Text](#)
59. Jilaihaw H, Doctor N, Kashif M, *et al.*: **Aortic annular sizing for transcatheter aortic valve replacement using cross-sectional 3-dimensional transesophageal echocardiography.** *J Am Coll Cardiol.* 2013; 61(9): 908–16.
[PubMed Abstract](#) | [Publisher Full Text](#)
60. Khaliq OK, Kodali SK, Paradis JM, *et al.*: **Aortic annular sizing using a novel 3-dimensional echocardiographic method: use and comparison with cardiac computed tomography.** *Circ Cardiovasc Imaging.* 2014; 7(1): 155–63.
[PubMed Abstract](#) | [Publisher Full Text](#)
61. Tamborini G, Fusini L, Gripari P, *et al.*: **Feasibility and accuracy of 3DTEE versus CT for the evaluation of aortic valve annulus to left main ostium distance before transcatheter aortic valve implantation.** *JACC Cardiovasc Imaging.* 2012; 5(6): 579–88.
[PubMed Abstract](#) | [Publisher Full Text](#)
62. **F** Smith LA, Dworakowski R, Bhan A, *et al.*: **Real-time three-dimensional transesophageal echocardiography adds value to transcatheter aortic valve implantation.** *J Am Soc Echocardiogr.* 2013; 26(4): 359–69.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
63. Pibarot P, Hahn RT, Weissman NJ, *et al.*: **Assessment of paravalvular regurgitation following TAVR: a proposal of unifying grading scheme.** *JACC Cardiovasc Imaging.* 2015; 8(3): 340–60.
[PubMed Abstract](#) | [Publisher Full Text](#)
64. Biner S, Perk G, Kar S, *et al.*: **Utility of combined two-dimensional and three-dimensional transesophageal imaging for catheter-based mitral valve clip repair of mitral regurgitation.** *J Am Soc Echocardiogr.* 2011; 24(6): 611–7.
[PubMed Abstract](#) | [Publisher Full Text](#)
65. **F** Altioek E, Becker M, Hamada S, *et al.*: **Optimized guidance of percutaneous edge-to-edge repair of the mitral valve using real-time 3-D transesophageal echocardiography.** *Clin Res Cardiol.* 2011; 100(8): 675–81.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
66. Guarracino F, Baldassarri R, Ferro B, *et al.*: **Transesophageal echocardiography during MitraClip® procedure.** *Anesth Analg.* 2014; 118(6): 1188–96.
[PubMed Abstract](#) | [Publisher Full Text](#)
67. Mor-Avi V, Lang RM, Badano LP, *et al.*: **Current and evolving echocardiographic techniques for the quantitative evaluation of cardiac mechanics: ASE/EAE consensus statement on methodology and indications endorsed by the Japanese Society of Echocardiography.** *J Am Soc Echocardiogr.* 2011; 24(3): 277–313.
[PubMed Abstract](#) | [Publisher Full Text](#)
68. Miyazaki S, Daimon M, Miyazaki T, *et al.*: **Global longitudinal strain in relation to the severity of aortic stenosis: a two-dimensional speckle-tracking study.** *Echocardiography.* 2011; 28(7): 703–8.
[PubMed Abstract](#) | [Publisher Full Text](#)
69. Ng ACT, Delgado V, Bertini M, *et al.*: **Alterations in multidirectional myocardial functions in patients with aortic stenosis and preserved ejection fraction: a two-dimensional speckle tracking analysis.** *Eur Heart J.* 2011; 32(12): 1542–50.
[PubMed Abstract](#) | [Publisher Full Text](#)
70. Sengupta PP, Tajik AJ, Chandrasekaran K, *et al.*: **Twist mechanics of the left ventricle: principles and application.** *JACC Cardiovasc Imaging.* 2008; 1(3): 366–76.
[PubMed Abstract](#) | [Publisher Full Text](#)
71. Bauer F, Eltchaninoff H, Tron C, *et al.*: **Acute improvement in global and regional left ventricular systolic function after percutaneous heart valve implantation in patients with symptomatic aortic stenosis.** *Circulation.* 2004; 110(11): 1473–6.
[PubMed Abstract](#) | [Publisher Full Text](#)
72. Delgado M, Ruiz M, Mesa D, *et al.*: **Early improvement of the regional and global ventricle function estimated by two-dimensional speckle tracking echocardiography after percutaneous aortic valve implantation speckle tracking after CoreValve implantation.** *Echocardiography.* 2013; 30(1): 37–44.
[PubMed Abstract](#) | [Publisher Full Text](#)
73. **F** Kamperidis V, Joyce E, Debonnaire P, *et al.*: **Left ventricular functional recovery and remodeling in low-flow low-gradient severe aortic stenosis after transcatheter aortic valve implantation.** *J Am Soc Echocardiogr.* 2014; 27(8): 817–25.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
74. **F** Carstensen HG, Larsen LH, Hassager C, *et al.*: **Association of ischemic heart disease to global and regional longitudinal strain in asymptomatic aortic stenosis.** *Int J Cardiovasc Imaging.* 2015; 31(3): 485–95.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
75. Ross J Jr, Braunwald E: **Aortic stenosis.** *Circulation.* 1968; 38(1 Suppl): 61–7.
[PubMed Abstract](#) | [Publisher Full Text](#)
76. **F** Carasso S, Mutlak D, Lessick J, *et al.*: **Symptoms in severe aortic stenosis are associated with decreased compensatory circumferential myocardial mechanics.** *J Am Soc Echocardiogr.* 2015; 28(2): 218–25.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
77. **F** Attias D, Macron L, Dreyfus J, *et al.*: **Relationship between longitudinal strain and symptomatic status in aortic stenosis.** *J Am Soc Echocardiogr.* 2013; 26(8): 868–74.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)

78. Logstrup BB, Andersen HR, Thuesen L, *et al.*: **Left ventricular global systolic longitudinal deformation and prognosis 1 year after femoral and apical transcatheter aortic valve implantation.** *J Am Soc Echocardiogr.* 2013; 26(3): 246–54.
[PubMed Abstract](#) | [Publisher Full Text](#)
79. **F** Kamperidis V, van Rosendaal PJ, Ng ACT, *et al.*: **Impact of flow and left ventricular strain on outcome of patients with preserved left ventricular ejection fraction and low gradient severe aortic stenosis undergoing aortic valve replacement.** *Am J Cardiol.* 2014; 114(12): 1875–81.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
80. **F** Dahou A, Bartko PE, Capoulade R, *et al.*: **Usefulness of global left ventricular longitudinal strain for risk stratification in low ejection fraction, low-gradient aortic stenosis: results from the multicenter True or Pseudo-Severe Aortic Stenosis study.** *Circ Cardiovasc Imaging.* 2015; 8(3): e002117.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
81. **F** Nagata Y, Takeuchi M, Wu VC, *et al.*: **Prognostic value of LV deformation parameters using 2D and 3D speckle-tracking echocardiography in asymptomatic patients with severe aortic stenosis and preserved LV ejection fraction.** *JACC Cardiovasc Imaging.* 2015; 8(3): 235–45.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
82. **F** Kusunose K, Goodman A, Parikh R, *et al.*: **Incremental prognostic value of left ventricular global longitudinal strain in patients with aortic stenosis and preserved ejection fraction.** *Circ Cardiovasc Imaging.* 2014; 7(6): 938–45.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
83. Di Salvo G, Rea A, Mormile A, *et al.*: **Usefulness of bidimensional strain imaging for predicting outcome in asymptomatic patients aged ≤ 16 years with isolated moderate to severe aortic regurgitation.** *Am J Cardiol.* 2012; 110(7): 1051–5.
[PubMed Abstract](#) | [Publisher Full Text](#)
84. Iida N, Seo Y, Ishizu T, *et al.*: **Transmural compensation of myocardial deformation to preserve left ventricular ejection performance in chronic aortic regurgitation.** *J Am Soc Echocardiogr.* 2012; 25(6): 620–8.
[PubMed Abstract](#) | [Publisher Full Text](#)
85. Olsen NT, Sogaard P, Larsson HB, *et al.*: **Speckle-tracking echocardiography for predicting outcome in chronic aortic regurgitation during conservative management and after surgery.** *JACC Cardiovasc Imaging.* 2011; 4(3): 223–30.
[PubMed Abstract](#) | [Publisher Full Text](#)
86. Onishi T, Kawai H, Tatsumi K, *et al.*: **Preoperative systolic strain rate predicts postoperative left ventricular dysfunction in patients with chronic aortic regurgitation.** *Circ Cardiovasc Imaging.* 2010; 3(2): 134–41.
[PubMed Abstract](#) | [Publisher Full Text](#)
87. Park SH, Yang YA, Kim KY, *et al.*: **Left Ventricular Strain as Predictor of Chronic Aortic Regurgitation.** *J Cardiovasc Ultrasound.* 2015; 23(2): 78–85.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
88. **F** Nishimura RA, Otto CM, Bonow RO, *et al.*: **2014 AHA/ACC guideline for the management of patients with valvular heart disease: executive summary: a report of the American College of Cardiology/American Heart Association Task Force on Practice Guidelines.** *J Am Coll Cardiol.* 2014; 63(22): 2438–88.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
89. de Isla LP, de Agustín A, Rodrigo JL, *et al.*: **Chronic mitral regurgitation: a pilot study to assess preoperative left ventricular contractile function using speckle-tracking echocardiography.** *J Am Soc Echocardiogr.* 2009; 22(7): 831–8.
[PubMed Abstract](#) | [Publisher Full Text](#)
90. **F** Witkowski TG, Thomas JD, Debonnaire PJMR, *et al.*: **Global longitudinal strain predicts left ventricular dysfunction after mitral valve repair.** *Eur Heart J Cardiovasc Imaging.* 2013; 14(1): 69–76.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
91. Pandis D, Sengupta PP, Castillo JG, *et al.*: **Assessment of longitudinal myocardial mechanics in patients with degenerative mitral valve regurgitation predicts postoperative worsening of left ventricular systolic function.** *J Am Soc Echocardiogr.* 2014; 27(6): 627–38.
[PubMed Abstract](#) | [Publisher Full Text](#)
92. Smith LA, Monaghan MJ: **Monitoring of procedures: peri-interventional echo assessment for transcatheter aortic valve implantation.** *Eur Heart J Cardiovasc Imaging.* 2013; 14(9): 840–50.
[PubMed Abstract](#) | [Publisher Full Text](#)
93. **F** Osman F, Steeds R: **Use of intra-cardiac ultrasound in the diagnosis of prosthetic valve malfunction.** *Eur J Echocardiogr.* 2007; 8(5): 392–4.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
94. Bartel T, Bonaros N, Edlinger M, *et al.*: **Intracardiac echo and reduced radioccontrast requirements during TAVR.** *JACC Cardiovasc Imaging.* 2014; 7(3): 319–20.
[PubMed Abstract](#) | [Publisher Full Text](#)
95. **F** Corti R, Biaggi P, Gaemperli O, *et al.*: **Integrated x-ray and echocardiography imaging for structural heart interventions.** *EuroIntervention.* 2013; 9(7): 863–9.
[PubMed Abstract](#) | **F1000 Recommendation**
96. **F** Gao G, Penney G, Ma Y, *et al.*: **Registration of 3D trans-esophageal echocardiography to X-ray fluoroscopy using image-based probe tracking.** *Med Image Anal.* 2012; 16(1): 38–49.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
97. **F** Kaiser M, John M, Heimann T, *et al.*: **2D/3D registration of TEE probe from two non-orthogonal C-arm directions.** *Med Image Comput Comput Assist Interv.* 2014; 17(Pt 1): 283–90.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
98. Quaife RA, Salcedo EE, Carroll JD: **Procedural guidance using advance imaging techniques for percutaneous edge-to-edge mitral valve repair.** *Curr Cardiol Rep.* 2014; 16(2): 452.
[PubMed Abstract](#) | [Publisher Full Text](#)

Open Peer Review

Current Peer Review Status:

Version 1

Reviewer Report 28 September 2015

<https://doi.org/10.5256/f1000research.6917.r10568>

© 2015 Shiota T. This is an open access peer review report distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Takahiro Shiota

Cardiac Noninvasive Laboratory, Cedars-Sinai Heart Institute, Los Angeles, CA, USA

This review by Dr. Hahn from Columbia University covers major areas of the topic (recent advances in echocardiography in valvular heart disease) and is well written. With this, readers can appreciate her daily clinical experiences with new modalities such as three dimensional color Doppler echocardiography. In addition, recent major publications, including her own, are well chosen for the educational purpose.

Competing Interests: No competing interests were disclosed.

I confirm that I have read this submission and believe that I have an appropriate level of expertise to confirm that it is of an acceptable scientific standard.

Reviewer Report 28 September 2015

<https://doi.org/10.5256/f1000research.6917.r10567>

© 2015 Bax J. This is an open access peer review report distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Jeroen J. Bax

Department of Cardiology, Leiden University Medical Center, Leiden, The Netherlands

Competing Interests: No competing interests were disclosed.

I confirm that I have read this submission and believe that I have an appropriate level of expertise to confirm that it is of an acceptable scientific standard.

Reviewer Report 28 September 2015

<https://doi.org/10.5256/f1000research.6917.r10566>

© 2015 Hung J. This is an open access peer review report distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Judy Hung

Massachusetts General Hospital, Boston, MA, USA

Competing Interests: No competing interests were disclosed.

I confirm that I have read this submission and believe that I have an appropriate level of expertise to confirm that it is of an acceptable scientific standard.

The benefits of publishing with F1000Research:

- Your article is published within days, with no editorial bias
- You can publish traditional articles, null/negative results, case reports, data notes and more
- The peer review process is transparent and collaborative
- Your article is indexed in PubMed after passing peer review
- Dedicated customer support at every stage

For pre-submission enquiries, contact research@f1000.com

F1000Research