REVIEW

Recent advances in echocardiography: strain and strain rate imaging [version 1; peer review: 3 approved]

Oana Mirea¹, Jurgen Duchenne¹, Jens-Uwe Voigt^{1,2}

¹Department of Cardiovascular Sciences, KU Leuven – University of Leuven, Herestraat, Leuven, Belgium

²Department of Cardiovascular Diseases, UZ Leuven – University Hospitals Leuven, Herestraat, Leuven, Belgium

V1 First published: 29 Apr 2016, 5(F1000 Faculty Rev):787
<https://doi.org/10.12688/f1000research.7228.1>

Latest published: 29 Apr 2016, 5(F1000 Faculty Rev):787
<https://doi.org/10.12688/f1000research.7228.1>

Abstract

Deformation imaging by echocardiography is a well-established research tool which has been gaining interest from clinical cardiologists since the introduction of speckle tracking. Post-processing of echo images to analyze deformation has become readily available at the fingertips of the user. New parameters such as global longitudinal strain have been shown to provide added diagnostic value, and ongoing efforts of the imaging societies and industry aimed at harmonizing methods will improve the technique further. This review focuses on recent advances in the field of echocardiographic strain and strain rate imaging, and provides an overview on its current and potential future clinical applications.

Keywords

echocardiography, strain rate imaging, echocardiographic strain, speckle tracking

Open Peer Review

Approval Status

	1	2	3
version 1 29 Apr 2016			

Faculty Reviews are review articles written by the prestigious Members of **Faculty Opinions**. The articles are commissioned and peer reviewed before publication to ensure that the final, published version is comprehensive and accessible. The reviewers who approved the final version are listed with their names and affiliations.

1. **Giovanni Di Salvo**, Alfaisal University, Riyadh, Saudi Arabia
2. **Tomasz Kukulski**, Zabrze Silesian Medical University, Zabrze, Poland
3. **Alan Fraser**, Cardiff University School of Medicine, Cardiff, UK

Any comments on the article can be found at the end of the article.

Corresponding author: Jens-Uwe Voigt (jens-uwe.voigt@uz.kuleuven.ac.be)

Competing interests: The authors declare that they have no competing interests.

Grant information: The author(s) declared that no grants were involved in supporting this work.

Copyright: © 2016 Mirea O *et al.* This is an open access article distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

How to cite this article: Mirea O, Duchenne J and Voigt JU. **Recent advances in echocardiography: strain and strain rate imaging [version 1; peer review: 3 approved]** F1000Research 2016, 5(F1000 Faculty Rev):787 <https://doi.org/10.12688/f1000research.7228.1>

First published: 29 Apr 2016, 5(F1000 Faculty Rev):787 <https://doi.org/10.12688/f1000research.7228.1>

Introduction

For decades, two-dimensional (2D) and Doppler echocardiography were the central pillars of evaluating left ventricular (LV) function. For purely practical reasons, many clinicians still resort to measurements of LV ejection fraction (EF) as well as the visual analysis of myocardial wall motion when they evaluate LV global and regional performance. However, these methods have a significant inter-observer variability as they depend on the skills and experience of the user. The undisputed need for simple, readily accessible and reliable methods for evaluating function has driven industry towards the development of semi- or fully- automated methods and post-processing tools for quantifying LV function.

Starting from first experiences with tissue Doppler, velocity imaging was followed by deformation imaging. But with the introduction of speckle tracking for analyzing images, quantitative analysis of myocardial function by deformation imaging had its breakthrough into clinical echocardiography. The recent recommendations released by the European Association of Cardiovascular Imaging (EACVI) and the American Society of Echocardiography (ASE) acknowledged the additional value of deformation measurements over traditional functional parameters, such as LV EF, and recommended the technique now for clinical use¹. This review will discuss the most recent developments in the field of strain imaging and the application of the method in the clinical setting.

What is strain?

Strain is defined as the fractional change in length of a myocardial segment relative to its baseline length, and it is expressed as a percentage. Strain rate is the temporal derivative of strain, and it provides information on the speed at which the deformation occurs. Strain is a vector and the complete description of the complex deformation of a piece of myocardium requires three normal and six shear strain components. For practical reasons, the normal strains which are preferred for clinical use are oriented along the coordinate system of the LV; they describe radial thickening and thinning as well as circumferential and longitudinal shortening and lengthening. Lengthening or thickening of the myocardium is represented by positive strain values, whereas negative values represent shortening or thinning. The most commonly used parameter is longitudinal strain, which can be expected to be around 20% in all regions of the LV².

Strain is ideally suited to quantify myocardial function regionally, but with the introduction of speckle tracking, a new parameter for global LV function assessment called “global strain” has been introduced. In the longitudinal direction, global longitudinal strain reflects the deformation along the entire LV wall which is visible in an apical image. The measurements from all three apical views are combined to give an average GLS value.

It must be noted that myocardial deformation is load-dependent. Therefore, strain and strain rate measurements must be interpreted considering ventricular wall thickness and shape as well as pre- and after-load.

How to measure strain?

Tissue Doppler imaging

Tissue Doppler Imaging (TDI) was the first method used for directly measuring myocardial deformation by echocardiography. Since a regional velocity gradient is analytically identical with the temporal derivative of a change in length, strain rate can be directly calculated from two velocity samples at a known distance apart. Integrating strain rate over time results in strain³. The method is well validated⁴ and has been shown to provide valuable data in a wide range of conditions⁵. It benefits particularly from the high frame rate of echocardiographic TDI and therefore is the method of choice in all situations where short-lived mechanical events and fast changes in deformation (e.g., in diastole) have to be measured.

Speckle-tracking echocardiography

Just over a decade ago, speckle-tracking echocardiography (STE) was proposed^{6,7} and validated^{8–10} as an alternative tool for measuring myocardial function. Observations in large patient populations showed encouraging results regarding its applicability^{8,11} in the clinical setting. The tracking algorithm identifies specific myocardial patterns (commonly named “speckles” or “features”) on conventional B-mode echocardiographic images and follows the motion of these patterns frame-by-frame. The potential to track the speckles in any direction within a 2D image allows the calculation of myocardial velocities, displacement, strain and strain rate in any given direction. This multidirectional tracking ability along with its angle independency¹² are often regarded as major advantages over TDI. However, STE has also been shown to depend significantly on good image quality and proper image geometry. Because speckle tracking is derived from grey-scale images which have lower frame rates than TDI, measurements of motion and deformation are most reliable for events that last longer, such as systole. Tracking-based measurements of velocity and rate of deformation, however, should only be attempted with caution. The fast and user-friendly (semi-)automated post-processing is the biggest advantage of the technique.

Three-dimensional speckle-tracking echocardiography

In theory, tracking can be performed in all three dimensions when three-dimensional (3D) echocardiographic image data are available. However, the problems of frame rate and image quality as explained for 2D tracking are potentiated in 3D, while the advantage of a fully three-dimensional deformation analysis remains a hypothesis. Therefore, 3D STE must currently be regarded as an experimental method¹³.

Definitions and conventions

To achieve reproducible measurements, the method of measuring strain needs to be defined and communicated together with the result. The definition comprises not only the interrogated deformation component (longitudinal, circumferential, or radial) but also the sampling position within the myocardium (e.g., endocardial, midwall, or full wall) and the temporal definition of the measured parameter, as conventions are lacking in this field and the definitions are dependent on the vendor of the analysis software. In

particular, definitions of timing can influence the measurement without being noticed by the user.

Myocardial deformation is a cyclical process, and the definition of when in this cycle the myocardium can be assumed to have reached its “baseline length” is completely arbitrary. However, it is extremely relevant for defining what “zero strain” is. In physiology lessons, we are taught to consider end-diastole as the reference point in the cardiac cycle, but this is a difficult definition as it requires the mitral valve closure to be measured. Therefore, most strain analysis softwares use surrogate parameters such as the R-peak of the QRS complex in the electrocardiogram. The time of the R-peak, however, can deviate considerably from the true mitral valve closure time, in particular when conduction delays are present¹⁴.

A similar question occurs when it comes to the definition of end-systole. In physiology, we use the time of aortic valve closure. In strain software, however, several vendors use the nadir of the global strain or volume curve as an approximation, since it can be calculated easily from the tracking data. Again, particularly in conduction delays or regional dysfunction, this surrogate can be very wrong¹⁴. Therefore, a good software should allow the measurement of aortic valve closure and its implementation in the strain analysis.

Figure 1 displays the impact of timing changes on strain measurements. Figure 2 shows an overview of commonly used definitions,

such as peak systolic strain, end-systolic strain, post-systolic strain (PSS), and peak strain. As can be seen, the definition of end-systole is of particular interest, as derived parameters such as post-systolic shortening directly depend on it.

Recent clinical applications of deformation imaging

In recent years, functional imaging based on speckle tracking has entered the clinical arena. Hundreds of publications have explored the potential of STE to improve the prognostic and diagnostic accuracy of echocardiography in a wide variety of cardiac pathologies associated with LV dysfunction.

Prognostic implications

In the general population, lower GLS was shown to be a powerful and independent predictor for vascular events such as stroke and myocardial infarction and for new onset of atrial fibrillation^{15,16}. GLS also demonstrated prognostic value for adverse outcomes in patients with heart failure^{17–19}, coronary artery disease (CAD)^{20–22}, valvular heart disease^{23,24}, and cardiomyopathies²⁵.

Furthermore, GLS proved to be a superior predictor of all-cause mortality when compared with LV EF and myocardial wall motion in patients with CAD^{26,27} or chronic kidney disease²⁸. Longitudinal strain measurements also showed encouraging results in identifying early LV impairment in patients undergoing chemotherapy^{29,30} and in subjects with chronic nephropathy³¹ or diabetes mellitus³².

Figure 1. End-systolic strain variability due to changes in timing definitions. **A)** Speckle tracking longitudinal strain in a dilated ventricle with left bundle branch block. The yellow and blue dot indicate the origin of the strain curves of the same colour in the lower panels, where in addition a dotted white curve represents global longitudinal strain. **(B)** Correct definition of end-diastole (ED) and end-systole (ES). **(C)** The definition of ES has been moved by + 4 frames (green arrow indicates initial ES; green dashed line indicates current ES). Note the impact on the measured systolic strain value. **(D)** Here, the definition of ED has been moved by + 4 frames (white arrow indicates initial ED; white dashed line indicates current ED). Also, in this case, the measurement of end-systolic strain is indirectly affected due to a shift in baseline.

Figure 2. Overview and definition of commonly used strain measurements. Peak systolic strain (SS) is always measured before aortic valve closure (AVC). End-systolic strain (ES) is measured on AVC. Post-systolic strain (PSS) peaks after AVC.

Coronary artery disease

Possibly the most valuable clinical application of strain measurements is the evaluation of regional dysfunction in patients with CAD. Early experimental validation against sonomicrometry showed that both Doppler-derived⁴ and STE-derived³³ longitudinal strain can detect the presence of ischemia. Additionally, encouraging data were reported with respect to the ability of strain measurements to predict the extent of the ischemic area³⁴ and to differentiate between transmural and non-transmural scar³⁵.

PSS, defined as the presence of regional myocardial shortening after aortic valve closure, is considered a hallmark for myocardial ischemia (Figure 3). The presence of PSS in areas with ischemic insult has been demonstrated in early studies by using TDI-derived measurements^{36,37}. A longer persistence of the PSS after an ischemic event was associated with more severe coronary obstruction³⁸. Nevertheless, it is important to consider that although PSS is a very sensitive marker of regional dysfunction, it is never specific for a certain pathology; PSS always needs to be interpreted in a clinical or pathophysiological context or both. PSS at rest can be a sign of ischemia, myocardial scar, or other conditions³⁶. PSS which occurs during stress echocardiography, however, is very likely to be caused by ischemia³⁸, so it can improve the accuracy of detecting CAD

during a dobutamine stress test³⁹. Measurements of longitudinal strain during dobutamine stress echocardiography for detecting CAD were reported to be feasible in 75–100%^{40,41}. Whether these numbers are realistic in a routine clinical scenario, however, remains doubtful.

LV functional dispersion and dyssynchrony

Strain imaging allows the assessment of asynchronous LV deformation (e.g., by measuring the time to peak strain). Haugaa *et al.* demonstrated in patients with CAD that abnormalities in synchronicity, referred to as mechanical dispersion and defined as the standard deviation of the time to peak regional shortening, could identify those with high risk for arrhythmias⁴². An example of larger mechanical dispersion due to the presence of ischemia when compared with normal myocardium is presented in Figure 4.

In past years, many indices have been suggested to identify potential responders to cardiac resynchronization therapy (CRT). Initial parameters were based on time-to-peak velocity measurements and could successfully detect dyssynchrony but failed to show added predictive value in prospective clinical trials^{43–48}. The reason can be found in the inability of velocity measurements to distinguish wall motion due to contraction and wall motion due to

Figure 3. Abnormal deformation in myocardial infarction. (A) Segmental longitudinal strain curves in an apical two-chamber view of a patient with inferior infarction. The bold yellow curve is derived from the infarcted inferobasal segment. Note the pronounced post-systolic shortening (arrow). (B) Cardiac magnetic resonance imaging with delayed enhancement of the same patient. A scar is present in the basal inferior region (arrows). Abbreviations: AVC, aortic valve closure; GS, global strain.

Figure 4. Mechanical dispersion (MD) of the segmental peak longitudinal strain. (A) Normal heart. All strain curves peak around aortic valve closure (AVC). (B) Infarct patient with inducible arrhythmias in the electrophysiology lab. Note the wide dispersion of the segmental peak strains. Abbreviations: GS, global strain.

tethering which prevents conclusive description of the sequence of wall activation. Later, deformation-based parameters have been suggested, focusing on the timing difference measured between dyssynchronous walls^{49–54}. Such parameters are potentially useful but only if they identify signatures in the myocardial deformation pattern which are specific for hearts amenable to CRT. More recently, a new index has been proposed which is based on the non-invasive assessment of regional myocardial work and which combines regional LV deformation with an estimate of LV pressure^{55,56}. It remains to be determined whether such parameters will prove to be superior to the more easy, direct visual or quantitative evaluation of the characteristic motion patterns of dyssynchronous hearts^{57–59}. Both the fast early systolic inward motion and the subsequent stretching of the septum (septal flash), as well as the rocking motion of the LV apex (apical rocking), can be directly assessed and have been shown to successfully predict CRT response in a clinical setting with both a qualitative and a quantitative approach⁶⁰.

Other recent applications

Recent research has demonstrated the complementary character of GLS and EF. Both parameters describe global LV function, but both do it in a different way. Although the parameters follow a linear fit of $EF = 3|GLS|$ in most situations⁶¹, GLS and EF may diverge depending on the underlying pathology, which could offer added diagnostic information. Several studies have shown that GLS is more sensitive to subtle changes in myocardial function which, for example, could be used in the follow-up of patients receiving chemotherapy^{62,63}. In hypertrophic pathology, GLS is frequently reduced while EF is still normal⁶⁴. A higher EF/GLS ratio was found to differentiate cardiac amyloidosis from other pathologies with increased LV wall thickness, such as hypertrophic cardiomyopathy⁶⁵.

Advances in the standardization of strain measurements

Establishing the reliability of STE is a prerequisite to its clinical implementation. Although STE has proved to have numerous advantages, such as large availability, high feasibility, and added clinical information, there are still debates regarding the potential discrepancy of strain measurements between vendors^{66,67}.

In recognition of this problem, the EACVI has initiated a task force together with the ASE and industry partners. The aims of this task force are to identify and minimize sources of variability between strain measurements and to standardize definitions for strain measurements^{3,68}. The first results of the initiative provide insights on different speckle tracking algorithms, current terminology, and specific technical issues such as image acquisition or timing of measurements³.

In a first comparative study between all major vendors of echocardiography machines and strain software, GLS measurements have been found to show considerable inter-vendor differences, whereas its reproducibility was consistently comparable to or even better

than that of conventional echocardiographic parameters⁶⁹. These findings imply that GLS measurements can be considered a reliable tool in the clinical routine as long as repetitive measurements are performed with the same equipment. On the contrary, the comparison of data obtained with different post-processing software should be avoided.

Future perspectives

The assessment of LV regional function by speckle tracking has not yet been intensively tested. In particular, vendor-specific differences in the tracking algorithms, such as the number and dimension of kernels, regional smoothing, underlying models of LV motion, or others, may account for more significant differences at a segmental level than at a global level.

Moreover, since regional function can be interpreted through various parameters, the most reproducible and robust for the definition of regional dysfunction remains to be determined. Given the potential value of regional strain measurements, the ongoing efforts of the EACVI/ASE Task Force focus on identifying reasons for the variability of regional strain parameters between vendor-specific software. Positive results would be a remarkable step forward in the process of endorsing regional strain measurements.

Conclusions

Echocardiographic deformation imaging has developed into an indispensable tool for the clinical assessment of a wide range of cardiac pathologies, and current guidelines recommend its use because of its feasibility and robustness. However, one should remain aware of the pitfalls which still exist and which are easily overlooked because of the user-friendly one-click-gives-it-all approach of most software solutions. Joint efforts of echocardiography associations and industry partners to resolve these pitfalls, to minimize inter-vendor differences, and to standardize measurement definitions will provide a solid valid basis for a widespread use of the technique in the clinical routine.

Abbreviations

2D, two-dimensional; 3D, three-dimensional; ASE, American Society of Echocardiography; CAD, coronary artery disease; CRT, cardiac resynchronization therapy; EACVI, European Association of Cardiovascular Imaging; EF, ejection fraction; GLS, global longitudinal strain; LV, left ventricular; PSS, post-systolic strain; STE, speckle-tracking echocardiography; TDI, tissue Doppler imaging.

Competing interests

The authors declare that they have no competing interests.

Grant information

The author(s) declared that no grants were involved in supporting this work.

References

1. Lang RM, Badano LP, Mor-Avi V, *et al.*: Recommendations for cardiac chamber quantification by echocardiography in adults: an update from the American Society of Echocardiography and the European Association of Cardiovascular Imaging. *Eur Heart J Cardiovasc Imaging*. 2015; 16(3): 233–70.
[PubMed Abstract](#) | [Publisher Full Text](#)
2. Yingchoncharoen T, Agarwal S, Popović ZB, *et al.*: Normal ranges of left ventricular strain: a meta-analysis. *J Am Soc Echocardiogr*. 2013; 26(2): 185–91.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
3. Voigt JU, Pedrizzetti G, Lysyansky P, *et al.*: Definitions for a common standard for 2D speckle tracking echocardiography: consensus document of the EACVI/ASE/Industry Task Force to standardize deformation imaging. *Eur Heart J Cardiovasc Imaging*. 2015; 16(1): 1–11.
[PubMed Abstract](#) | [Publisher Full Text](#)
4. Urheim S, Edvardsen T, Torp H, *et al.*: Myocardial strain by Doppler echocardiography. Validation of a new method to quantify regional myocardial function. *Circulation*. 2000; 102(10): 1158–64.
[PubMed Abstract](#) | [Publisher Full Text](#)
5. Ho CY, Solomon SD: A clinician's guide to tissue Doppler imaging. *Circulation*. 2006; 113(10): e396–8.
[PubMed Abstract](#) | [Publisher Full Text](#)
6. Toyoda T, Baba H, Akasaka T, *et al.*: Assessment of regional myocardial strain by a novel automated tracking system from digital image files. *J Am Soc Echocardiogr*. 2004; 17(12): 1234–8.
[PubMed Abstract](#) | [Publisher Full Text](#)
7. Leitman M, Lysyansky P, Sidenko S, *et al.*: Two-dimensional strain—a novel software for real-time quantitative echocardiographic assessment of myocardial function. *J Am Soc Echocardiogr*. 2004; 17(10): 1021–9.
[PubMed Abstract](#) | [Publisher Full Text](#)
8. Amundsen BH, Helle-Valle T, Edvardsen T, *et al.*: Noninvasive myocardial strain measurement by speckle tracking echocardiography: validation against sonomicrometry and tagged magnetic resonance imaging. *J Am Coll Cardiol*. 2006; 47(4): 789–93.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
9. Korinek J, Wang J, Sengupta PP, *et al.*: Two-dimensional strain—a Doppler-independent ultrasound method for quantitation of regional deformation: validation *in vitro* and *in vivo*. *J Am Soc Echocardiogr*. 2005; 18(12): 1247–53.
[PubMed Abstract](#) | [Publisher Full Text](#)
10. Edvardsen T, Gerber BL, Garot J, *et al.*: Quantitative assessment of intrinsic regional myocardial deformation by Doppler strain rate echocardiography in humans: validation against three-dimensional tagged magnetic resonance imaging. *Circulation*. 2002; 106(1): 50–6.
[PubMed Abstract](#) | [Publisher Full Text](#)
11. D'hooge J, Konofagou E, Jamal F, *et al.*: Two-dimensional ultrasonic strain rate measurement of the human heart *in vivo*. *IEEE Trans Ultrason Ferroelectr Freq Control*. 2002; 49(2): 281–6.
[PubMed Abstract](#) | [Publisher Full Text](#)
12. Forsha D, Risum N, Rajagopal S, *et al.*: The influence of angle of insonation and target depth on speckle-tracking strain. *J Am Soc Echocardiogr*. 2015; 28(5): 580–6.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
13. Badano LP, Cucchini U, Muraru D, *et al.*: Use of three-dimensional speckle tracking to assess left ventricular myocardial mechanics: inter-vendor consistency and reproducibility of strain measurements. *Eur Heart J Cardiovasc Imaging*. 2013; 14(3): 285–93.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
14. Mada RO, Lysyansky P, Daraban AM, *et al.*: How to define end-diastole and end-systole?: Impact of timing on strain measurements. *JACC Cardiovasc Imaging*. 2015; 8(2): 148–57.
[PubMed Abstract](#) | [Publisher Full Text](#)
15. Russo C, Jin Z, Elkind MS, *et al.*: Prevalence and prognostic value of subclinical left ventricular systolic dysfunction by global longitudinal strain in a community-based cohort. *Eur J Heart Fail*. 2014; 16(12): 1301–9.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | **F1000 Recommendation**
16. Russo C, Jin Z, Sera F, *et al.*: Left Ventricular Systolic Dysfunction by Longitudinal Strain Is an Independent Predictor of Incident Atrial Fibrillation: A Community-Based Cohort Study. *Circ Cardiovasc Imaging*. 2015; 8(8): e003520.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | **F1000 Recommendation**
17. Motoki H, Borowski AG, Shrestha K, *et al.*: Incremental prognostic value of assessing left ventricular myocardial mechanics in patients with chronic systolic heart failure. *J Am Coll Cardiol*. 2012; 60(20): 2074–81.
[PubMed Abstract](#) | [Publisher Full Text](#)
18. Zhang KW, French B, May Khan A, *et al.*: Strain improves risk prediction beyond ejection fraction in chronic systolic heart failure. *J Am Heart Assoc*. 2014; 3(1): e000550.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
19. Iacoviello M, Puzzovivo A, Guida P, *et al.*: Independent role of left ventricular global longitudinal strain in predicting prognosis of chronic heart failure patients. *Echocardiography*. 2013; 30(7): 803–11.
[PubMed Abstract](#) | [Publisher Full Text](#)
20. Ersbøll M, Valeur N, Mogensen UM, *et al.*: Prediction of all-cause mortality and heart failure admissions from global left ventricular longitudinal strain in patients with acute myocardial infarction and preserved left ventricular ejection fraction. *J Am Coll Cardiol*. 2013; 61(23): 2365–73.
[PubMed Abstract](#) | [Publisher Full Text](#)
21. Munk K, Andersen NH, Terkelsen CJ, *et al.*: Global left ventricular longitudinal systolic strain for early risk assessment in patients with acute myocardial infarction treated with primary percutaneous intervention. *J Am Soc Echocardiogr*. 2012; 25(6): 644–51.
[PubMed Abstract](#) | [Publisher Full Text](#)
22. Bertini M, Ng AC, Antoni ML, *et al.*: Global longitudinal strain predicts long-term survival in patients with chronic ischemic cardiomyopathy. *Circ Cardiovasc Imaging*. 2012; 5(3): 383–91.
[PubMed Abstract](#) | [Publisher Full Text](#)
23. Lancellotti P, Cosyns B, Zacharakis D, *et al.*: Importance of left ventricular longitudinal function and functional reserve in patients with degenerative mitral regurgitation: assessment by two-dimensional speckle tracking. *J Am Soc Echocardiogr*. 2008; 21(12): 1331–6.
[PubMed Abstract](#) | [Publisher Full Text](#)
24. Dahl JS, Videbæk L, Poulsen MK, *et al.*: Global strain in severe aortic valve stenosis: relation to clinical outcome after aortic valve replacement. *Circ Cardiovasc Imaging*. 2012; 5(5): 613–20.
[PubMed Abstract](#) | [Publisher Full Text](#)
25. Buss SJ, Emami M, Mereles D, *et al.*: Longitudinal left ventricular function for prediction of survival in systemic light-chain amyloidosis: incremental value compared with clinical and biochemical markers. *J Am Coll Cardiol*. 2012; 60(12): 1067–76.
[PubMed Abstract](#) | [Publisher Full Text](#)
26. Stanton T, Leano R, Marwick TH: Prediction of all-cause mortality from global longitudinal speckle strain: comparison with ejection fraction and wall motion scoring. *Circ Cardiovasc Imaging*. 2009; 2(5): 356–64.
[PubMed Abstract](#) | [Publisher Full Text](#)
27. Antoni ML, Mollema SA, Delgado V, *et al.*: Prognostic importance of strain and strain rate after acute myocardial infarction. *Eur Heart J*. 2010; 31(13): 1640–7.
[PubMed Abstract](#) | [Publisher Full Text](#)
28. Krishnasamy R, Isbel NM, Hawley CM, *et al.*: Left Ventricular Global Longitudinal Strain (GLS) Is a Superior Predictor of All-Cause and Cardiovascular Mortality When Compared to Ejection Fraction in Advanced Chronic Kidney Disease. *PLoS One*. 2015; 10(5): e0127044.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
29. Stoodley PW, Richards DA, Hui R, *et al.*: Two-dimensional myocardial strain imaging detects changes in left ventricular systolic function immediately after anthracycline chemotherapy. *Eur J Echocardiogr*. 2011; 12(12): 945–52.
[PubMed Abstract](#) | [Publisher Full Text](#)
30. Sawaya H, Sebag IA, Plana JC, *et al.*: Early detection and prediction of cardiotoxicity in chemotherapy-treated patients. *Am J Cardiol*. 2011; 107(9): 1375–80.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
31. Panoulas VF, Sulemane S, Konstantinou K, *et al.*: Early detection of subclinical left ventricular myocardial dysfunction in patients with chronic kidney disease. *Eur Heart J Cardiovasc Imaging*. 2015; 16(5): 539–48.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
32. Wang Q, Gao Y, Tan K, *et al.*: Subclinical impairment of left ventricular function in diabetic patients with or without obesity: A study based on three-dimensional speckle tracking echocardiography. *Herz*. 2015; 40(Suppl 3): 260–8.
[PubMed Abstract](#) | [Publisher Full Text](#) | **F1000 Recommendation**
33. Aarsaether E, Rösner A, Straumbotn E, *et al.*: Peak longitudinal strain most accurately reflects myocardial segmental viability following acute myocardial infarction - an experimental study in open-chest pigs. *Cardiovasc Ultrasound*. 2012; 10: 23.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | **F1000 Recommendation**
34. Choi JO, Cho SW, Song YB, *et al.*: Longitudinal 2D strain at rest predicts the presence of left main and three vessel coronary artery disease in patients without regional wall motion abnormality. *Eur J Echocardiogr*. 2009; 10(5): 695–701.
[PubMed Abstract](#) | [Publisher Full Text](#)
35. Roes SD, Mollema SA, Lamb HJ, *et al.*: Validation of echocardiographic two-dimensional speckle tracking longitudinal strain imaging for viability assessment in patients with chronic ischemic left ventricular dysfunction and comparison with contrast-enhanced magnetic resonance imaging. *Am J Cardiol*. 2009; 104(3): 312–7.
[PubMed Abstract](#) | [Publisher Full Text](#)
36. Voigt JU, Lindenmeier G, Exner B, *et al.*: Incidence and characteristics of segmental postsystolic longitudinal shortening in normal, acutely ischemic, and scarred myocardium. *J Am Soc Echocardiogr*. 2003; 16(5): 415–23.
[PubMed Abstract](#) | [Publisher Full Text](#)
37. Pislaru C, Anagnostopoulos PC, Seward JB, *et al.*: Higher myocardial strain rates

- during isovolumic relaxation phase than during ejection characterize acutely ischemic myocardium. *J Am Coll Cardiol*. 2002; 40(8): 1487–94.
[PubMed Abstract](#) | [Publisher Full Text](#)
38. Voigt JU, Exner B, Schmiedehausen K, *et al.*: Strain-rate imaging during dobutamine stress echocardiography provides objective evidence of inducible ischemia. *Circulation*. 2003; 107(16): 2120–6.
[PubMed Abstract](#) | [Publisher Full Text](#)
 39. Asanuma T, Uranishi A, Masuda K, *et al.*: Assessment of myocardial ischemic memory using persistence of post-systolic thickening after recovery from ischemia. *JACC Cardiovasc Imaging*. 2009; 2(11): 1253–61.
[PubMed Abstract](#) | [Publisher Full Text](#)
 40. F Wierzbowska-Drabik K, Hamala P, Roszczyk N, *et al.*: Feasibility and correlation of standard 2D speckle tracking echocardiography and automated function imaging derived parameters of left ventricular function during dobutamine stress test. *Int J Cardiovasc Imaging*. 2014; 30(4): 729–37.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | [F1000 Recommendation](#)
 41. F Yamada A, Luis SA, Sathianathan D, *et al.*: Reproducibility of regional and global longitudinal strains derived from two-dimensional speckle-tracking and Doppler tissue imaging between expert and novice readers during quantitative dobutamine stress echocardiography. *J Am Soc Echocardiogr*. 2014; 27(8): 880–7.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 42. Haugaa KH, Smedsrud MK, Steen T, *et al.*: Mechanical dispersion assessed by myocardial strain in patients after myocardial infarction for risk prediction of ventricular arrhythmia. *JACC Cardiovasc Imaging*. 2010; 3(3): 247–56.
[PubMed Abstract](#) | [Publisher Full Text](#)
 43. Yu CM, Chau E, Sanderson JE, *et al.*: Tissue Doppler echocardiographic evidence of reverse remodeling and improved synchronicity by simultaneously delaying regional contraction after biventricular pacing therapy in heart failure. *Circulation*. 2002; 105(4): 438–45.
[PubMed Abstract](#) | [Publisher Full Text](#)
 44. Bax JJ, Bleeker GB, Marwick TH, *et al.*: Left ventricular dyssynchrony predicts response and prognosis after cardiac resynchronization therapy. *J Am Coll Cardiol*. 2004; 44(9): 1834–40.
[PubMed Abstract](#) | [Publisher Full Text](#)
 45. Yu CM, Gorcsan J 3rd, Bleeker GB, *et al.*: Usefulness of tissue Doppler velocity and strain dyssynchrony for predicting left ventricular reverse remodeling response after cardiac resynchronization therapy. *Am J Cardiol*. 2007; 100(8): 1263–70.
[PubMed Abstract](#) | [Publisher Full Text](#)
 46. F Chung ES, Leon AR, Tavazzi L, *et al.*: Results of the Predictors of Response to CRT (PROSPECT) trial. *Circulation*. 2008; 117(20): 2608–16.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 47. Miyazaki C, Redfield MM, Powell BD, *et al.*: Dyssynchrony indices to predict response to cardiac resynchronization therapy: a comprehensive prospective single-center study. *Circ Heart Fail*. 2010; 3(5): 565–73.
[PubMed Abstract](#) | [Publisher Full Text](#)
 48. Breithardt OA, Stellbrink C, Herbots L, *et al.*: Cardiac resynchronization therapy can reverse abnormal myocardial strain distribution in patients with heart failure and left bundle branch block. *J Am Coll Cardiol*. 2003; 42(3): 486–94.
[PubMed Abstract](#) | [Publisher Full Text](#)
 49. Reant P, Zaroui A, Donal E, *et al.*: Identification and characterization of super-responders after cardiac resynchronization therapy. *Am J Cardiol*. 2010; 105(9): 1327–35.
[PubMed Abstract](#) | [Publisher Full Text](#)
 50. Mele D, Toselli T, Capasso F, *et al.*: Comparison of myocardial deformation and velocity dyssynchrony for identification of responders to cardiac resynchronization therapy. *Eur J Heart Fail*. 2009; 11(4): 391–9.
[PubMed Abstract](#) | [Publisher Full Text](#)
 51. Lim P, Donal E, Lafitte S, *et al.*: Multicentre study using strain delay index for predicting response to cardiac resynchronization therapy (MUSIC study). *Eur J Heart Fail*. 2011; 13(9): 984–91.
[PubMed Abstract](#) | [Publisher Full Text](#)
 52. Gorcsan J 3rd, Oyenuga O, Habib PJ, *et al.*: Relationship of echocardiographic dyssynchrony to long-term survival after cardiac resynchronization therapy. *Circulation*. 2010; 122(19): 1910–8.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
 53. Tanaka H, Nesser HJ, Buck T, *et al.*: Dyssynchrony by speckle-tracking echocardiography and response to cardiac resynchronization therapy: results of the Speckle Tracking and Resynchronization (STAR) study. *Eur Heart J*. 2010; 31(14): 1690–700.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
 54. Delgado V, Ypenburg C, van Bommel RJ, *et al.*: Assessment of left ventricular dyssynchrony by speckle tracking strain imaging comparison between longitudinal, circumferential, and radial strain in cardiac resynchronization therapy. *J Am Coll Cardiol*. 2008; 51(20): 1944–52.
[PubMed Abstract](#) | [Publisher Full Text](#)
 55. F Russell K, Eriksen M, Aaberge L, *et al.*: A novel clinical method for quantification of regional left ventricular pressure-strain loop area: a non-invasive index of myocardial work. *Eur Heart J*. 2012; 33(6): 724–33.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | [F1000 Recommendation](#)
 56. F Russell K, Eriksen M, Aaberge L, *et al.*: Assessment of wasted myocardial work: a novel method to quantify energy loss due to uncoordinated left ventricular contractions. *Am J Physiol Heart Circ Physiol*. 2013; 305(7): H996–1003.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 57. Leenders GE, Lumens J, Cramer MJ, *et al.*: Septal deformation patterns delineate mechanical dyssynchrony and regional differences in contractility: analysis of patient data using a computer model. *Circ Heart Fail*. 2012; 5(1): 87–96.
[PubMed Abstract](#) | [Publisher Full Text](#)
 58. Szulik M, Tillekaerts M, Vangeel V, *et al.*: Assessment of apical rocking: a new, integrative approach for selection of candidates for cardiac resynchronization therapy. *Eur J Echocardiogr*. 2010; 11(10): 863–9.
[PubMed Abstract](#) | [Publisher Full Text](#)
 59. Stankovic I, Aaronson M, Smith HJ, *et al.*: Dynamic relationship of left-ventricular dyssynchrony and contractile reserve in patients undergoing cardiac resynchronization therapy. *Eur Heart J*. 2014; 35(1): 48–55.
[PubMed Abstract](#) | [Publisher Full Text](#)
 60. Stankovic I, Prinz C, Ciarka A, *et al.*: Relationship of visually assessed apical rocking and septal flash to response and long-term survival following cardiac resynchronization therapy (PREDICT-CRT). *Eur Heart J Cardiovasc Imaging*. 2016; 17(3): 262–9.
[PubMed Abstract](#) | [Publisher Full Text](#)
 61. F Pedrizzetti G, Mangual J, Tonti G: On the geometrical relationship between global longitudinal strain and ejection fraction in the evaluation of cardiac contraction. *J Biomech*. 2014; 47(3): 746–9.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 62. F Thavandiranathan P, Poulin F, Lim KD, *et al.*: Use of myocardial strain imaging by echocardiography for the early detection of cardiotoxicity in patients during and after cancer chemotherapy: a systematic review. *J Am Coll Cardiol*. 2014; 63(25 Pt A): 2751–68.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 63. Florescu M, Magda LS, Enescu OA, *et al.*: Early detection of epirubicin-induced cardiotoxicity in patients with breast cancer. *J Am Soc Echocardiogr*. 2014; 27(1): 83–92.
[PubMed Abstract](#) | [Publisher Full Text](#)
 64. Ozawa K, Funabashi N, Takaoka H, *et al.*: Characteristic myocardial strain identified in hypertrophic cardiomyopathy subjects with preserved left ventricular ejection fraction using a novel multi-layer transthoracic echocardiography technique. *Int J Cardiol*. 2015; 184: 237–43.
[PubMed Abstract](#) | [Publisher Full Text](#)
 65. Pagourelas ED, Duchenne J, Mirea O, *et al.*: The Relation of Ejection Fraction and Global Longitudinal Strain in Amyloidosis: Implications for Differential Diagnosis. *JACC Cardiovasc Imaging*. 2016; pii: S1936-878X(15)00981-X.
[PubMed Abstract](#) | [Publisher Full Text](#)
 66. Risum N, Ali S, Olsen NT, *et al.*: Variability of global left ventricular deformation analysis using vendor dependent and independent two-dimensional speckle-tracking software in adults. *J Am Soc Echocardiogr*. 2012; 25(11): 1195–203.
[PubMed Abstract](#) | [Publisher Full Text](#)
 67. Bansal M, Cho GY, Chan J, *et al.*: Feasibility and accuracy of different techniques of two-dimensional speckle based strain and validation with harmonic phase magnetic resonance imaging. *J Am Soc Echocardiogr*. 2008; 21(12): 1318–25.
[PubMed Abstract](#) | [Publisher Full Text](#)
 68. Thomas JD, Badano LP: EACVI-ASE-industry initiative to standardize deformation imaging: a brief update from the co-chairs. *Eur Heart J Cardiovasc Imaging*. 2013; 14(11): 1039–40.
[PubMed Abstract](#) | [Publisher Full Text](#)
 69. F Farsalinos KE, Daraban AM, Ünlü S, *et al.*: Head-to-Head Comparison of Global Longitudinal Strain Measurements among Nine Different Vendors: The EACVI/ASE Inter-Vendor Comparison Study. *J Am Soc Echocardiogr*. 2015; 28(10): 1171–1181, e2.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)

Open Peer Review

Current Peer Review Status:

Editorial Note on the Review Process

Faculty Reviews are review articles written by the prestigious Members of **Faculty Opinions**. The articles are commissioned and peer reviewed before publication to ensure that the final, published version is comprehensive and accessible. The reviewers who approved the final version are listed with their names and affiliations.

The reviewers who approved this article are:

Version 1

1. **Alan Fraser**

Institute of Molecular & Experimental Medicine, Cardiff University School of Medicine, Cardiff, UK

Competing Interests: No competing interests were disclosed.

2. **Tomasz Kukulski**

Silesian Center for Heart Diseases, Zabrze Silesian Medical University, Zabrze, Poland

Competing Interests: No competing interests were disclosed.

3. **Giovanni Di Salvo**

King Faisal Specialist Hospital and Research Centre, Heart Centre, Alfaisal University, Riyadh, Saudi Arabia

Competing Interests: No competing interests were disclosed.

The benefits of publishing with F1000Research:

- Your article is published within days, with no editorial bias
- You can publish traditional articles, null/negative results, case reports, data notes and more
- The peer review process is transparent and collaborative
- Your article is indexed in PubMed after passing peer review
- Dedicated customer support at every stage

For pre-submission enquiries, contact research@f1000.com

F1000Research