REVIEW

Recent trends in airway management [version 1; peer review: 2 approved]

Joelle Karlik, Michael Aziz

Oregon Health & Science University, Portland, OR, USA

V1 First published: 17 Feb 2017, 6(F1000 Faculty Rev):159
<https://doi.org/10.12688/f1000research.10311.1>Latest published: 17 Feb 2017, 6(F1000 Faculty Rev):159
<https://doi.org/10.12688/f1000research.10311.1>

Abstract

Tracheal intubation remains a life-saving procedure that is typically not difficult for experienced providers in routine conditions. Unfortunately, difficult intubation remains challenging to predict and intubation conditions may make the event life threatening. Recent technological advances aim to further improve the ease, speed, safety, and success of intubation but have not been fully investigated. Video laryngoscopy, though proven effective in the difficult airway, may result in different intubation success rates in various settings and in different providers' hands. The rescue surgical airway remains a rarely used but critical skill, and research continues to investigate optimal techniques. This review highlights some of the new thoughts and research on these important topics.

Keywords

tracheal intubation, video laryngoscopy, surgical airway, apneic oxygenation, awake intubation

Open Peer Review

Approval Status

	1	2
version 1 17 Feb 2017		

Faculty Reviews are review articles written by the prestigious Members of **Faculty Opinions**. The articles are commissioned and peer reviewed before publication to ensure that the final, published version is comprehensive and accessible. The reviewers who approved the final version are listed with their names and affiliations.

1. **Jarrod Mosier**, University of Arizona, Tuscon, USA
2. **Narasimhan Jagannathan**, Northwestern University, Chicago, USA

Any comments on the article can be found at the end of the article.

Corresponding author: Michael Aziz (azizm@ohsu.edu)

Competing interests: The authors declare that they have no competing interests.

Grant information: The author(s) declared that no grants were involved in supporting this work.

Copyright: © 2017 Karlik J and Aziz M. This is an open access article distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

How to cite this article: Karlik J and Aziz M. **Recent trends in airway management [version 1; peer review: 2 approved]**
F1000Research 2017, 6(F1000 Faculty Rev):159 <https://doi.org/10.12688/f1000research.10311.1>

First published: 17 Feb 2017, 6(F1000 Faculty Rev):159 <https://doi.org/10.12688/f1000research.10311.1>

Introduction

Airway management remains a difficult skill to master, requiring hands-on training with human patients and extensive clinical experience. Fortunately, most intubations are not challenging in experienced hands. Current prediction models to anticipate the difficult intubation are poor and have limited application to new tools other than the traditional direct laryngoscopy (DL). Recent reviews of cannot-intubate cannot-oxygenate (CICO) events have continued to highlight the low rate of success for rescue despite the application by experienced anesthesiologists and surgeons¹. Fortunately, recent advances in airway management have facilitated improved intubation conditions by augmenting oxygenation during laryngoscopy and providing indirect video laryngeal views. This review will briefly discuss recent research in these areas, highlighting that all devices may not be appropriate in all health-care settings. Further emphasis is placed on recent research and recommendations on the definitive surgical airway in the CICO scenario.

Apneic oxygenation

Apneic oxygenation is not a new concept, but new research has begun to emphasize its potential value during airway management. Oxygenation via standard nasal cannula can be limited by patient comfort and maximum flow rates. A recent study demonstrated that high-flow nasal cannula at 15 mL/min does not increase time to desaturation in the critically ill population². These populations with increased oxygen consumption and large shunts may require higher delivery of apneic oxygenation than standard nasal cannula can provide.

Alternative devices offer higher concentrations of inspired oxygenation as well as increased comfort, including humidification. Various techniques for this enhanced apneic oxygenation include buccal RAE (Ring-Adair-Elwyn) tubes³, transnasal humidified rapid-insufflation ventilatory exchange (THRIVE) devices^{4,5}, and dual-use laryngoscopes⁶. These techniques show potential to substantially prolong apnea time. The THRIVE was recently found to be as effective as face-mask pre-oxygenation at maintaining oxygenation during rapid sequence induction⁷.

Video laryngoscopy

Video laryngoscopy (VL) has been shown to increase the rate of first-attempt success in patients with the predicted difficult airway in operating rooms^{8–10}. Accordingly, recent literature has focused on extending the use of VL to various settings, including pre-hospital, emergency department, and intensive care unit (ICU). These settings often present a high-stress scenario with an actively decompensating patient with widely varying provider experience. Current data suggest that the success seen with VL may not automatically translate to other settings outside the operating room.

The pre-hospital setting

VL has clear benefits and disadvantages in the pre-hospital setting. VL may have a special role for intubations in the field where patients are challenged with spine immobilization precautions, suboptimal positioning, and altered facial anatomy or trauma. However, secretions, blood, and vomitus may obscure VL cameras, and camera lenses are difficult to clear of contamination.

In addition, first responders may have different training and comfort levels with new devices.

The Airtraq (Prodol, Vizcaya, Spain), C-MAC (Karl Storz, Tuttlingen, Germany), and GlideScope (Verathon, Burnaby, BC, Canada) devices have been directly compared with DL in the pre-hospital emergency intubation setting. Initial observational studies have shown an association with decreased number of intubation attempts when using a GlideScope versus DL¹¹. A retrospective study in a pre-hospital paramedic helicopter system observed increased first-pass success and a decrease in overall attempts with C-MAC laryngoscopy¹². A prospective observational study of out-of-hospital intubations showed improved first-pass success with rapid sequence intubations but did not significantly change overall first-pass success when compared with DL¹³.

Further randomized research has questioned these benefits. Two prospective randomized trials of pre-hospital VL observed decreased rates for first-pass success when compared with DL^{14,15}. In another study, VL was significantly more successful in patients with a grade II–IV DL view, underlying the importance and success of indirect VL in patients with difficult DL views¹⁶.

VL may have a unique role in the pre-hospital setting where certain patient and environmental factors limit an adequate DL view¹⁷. In contrast, pre-hospital practitioners have a wide variety of training, skill, and experience that may limit their comfort with new devices. Given the challenges seen with indirect VL in the pre-hospital setting, we believe that devices with both DL and VL capabilities may be best in these environments, but this hypothesis is currently untested in the literature.

The emergency department setting

VL has also been studied in the emergency department setting where experience and a controlled hospital setting differ from pre-hospital emergency airways. Observational studies have demonstrated higher first-attempt success rates with VL over DL in the emergency department^{18,19}. However, a prospective randomized study showed no significant increase in first-pass success with C-MAC versus DL blades when used by senior residents in emergency resident training²⁰. A similar study of more experienced anesthesia providers showed no difference in first-pass success despite improved glottic views²¹. Overall, VL may not change initial intubation success in the emergency department but may be beneficial for subsequent attempts, but prospective randomized data are limited.

The intensive care setting

The intensive care setting often presents urgent or emergent airways in an actively decompensating patient. Observational data demonstrate a higher intubation success rate for VL over DL^{22,23}. Furthermore, in prospective randomized trials, the rate of successful first intubation was significantly higher with GlideScope VL than with DL in the hands of fellow-trained physicians, a finding also found in a study of resident physicians^{24,25}. However, a different trial with fellow physicians showed no significant differences in first-pass intubation success comparing

a variety of VL and DL devices²⁶. A recent multicenter study including the McGrath MAC video laryngoscope (Medtronic, Minneapolis, MN, USA) versus Macintosh laryngoscope (MACMAN) trial clarified optimal laryngoscopy technique in the ICU setting, but results again suggest that there is no difference in success rate between VL and DL in the ICU environment²⁷. The utility of VL in this environment, like that in other emergency environments, may be limited by the presence of blood in the airway²⁸.

Cardiopulmonary resuscitation

Intubation during cardiopulmonary resuscitation (CPR) provides definitive airway control and essential information about ongoing resuscitation via expired carbon dioxide. Unfortunately, successful intubation is often hampered by patient movement due to chest compressions. During CPR events, there was no difference in successful intubations or speed of intubation when prospectively comparing the GlideScope VL with DL²⁹. However, interruptions in chest compressions were minimized with VL in two different studies^{29,30}. In addition, VL was noted to have higher success rates when used by novice physicians during ongoing CPR³⁰.

Awake intubation

Awake intubations present a high-stress challenge with an anticipated difficult airway with a need for smooth, comfortable, and rapid intubation. In contrast to the scenarios discussed above, awake intubations are typically performed by providers with specific training in awake airway management techniques. Flexible bronchoscopic techniques are traditionally used for awake intubations because of their flexibility and comfort, but recent studies have investigated VL as an alternative technique.

A randomized trial of oral awake intubations with the flexible bronchoscope versus McGrath Series 5 VL (Teleflex, Morrisville, NC, USA) found no change in time to intubation or overall intubation success³¹. Average patient comfort scores after topical and regional anesthesia and sedation were similar for the two techniques. However, a high dropout rate in the VL group due to failed regional transtracheal block may limit interpretation of this study. Another randomized trial compared flexible bronchoscopic intubation with C-MAC D-blade VL for nasal awake intubations under topical anesthesia. It also showed no difference in the success of intubations³². These findings were validated in healthy volunteers in the awake upright (sitting) position^{33,34}. GlideScope VL was compared with fiber-optic techniques for the awake intubation of morbidly obese patients, a population with a higher rate of difficulty airway, with similar rates of successful intubation on first attempt³⁵. These cumulative findings suggest that VL may be a useful alternative to flexible bronchoscopic intubation in select patient populations with anticipated difficult airways.

Surgical airway

Scalpel versus cannula

The CICO scenario requiring surgical airway is a rare and stressful occurrence. Despite ongoing research and education in this area, the overall success of anesthesiologist-placed emergency surgical airway appears to be very low. The 4th National Audit Project (NAP4) recorded surgical airway management across the United Kingdom during a 1-year period. Most surgical airways were

performed by a surgeon with 100% success¹. In contrast, only 9 out of 25 (36%) anesthesiologists were successful in rescuing the airway with a surgical technique. Review of these cases underlined the importance of a simple rapid technique with an optimal chance of success in unpracticed and stressed providers.

Guidelines are currently divided between two techniques: scalpel or cannula airways. The 2015 Difficult Airway Society guidelines strongly favor scalpel techniques over cannula techniques, a change from the 2004 guidelines³⁶. In contrast, cannula techniques such as percutaneous cricothyrotomies are still included in similar guidelines, including the Australian and New Zealand College of Anaesthetists (ANZCA) Guidelines on Equipment to Manage a Difficult Airway During Anaesthesia³⁷.

Scalpel airways are thought to be faster, simpler, and overall more successful than cannula techniques. In the NAP4, 100% of first-choice surgical cricothyrotomies (3 out of 3) were successful, but the low number of first-choice scalpel airways may limit study interpretation¹. An observational study of success rates of trauma intubation and airway rescue techniques in both anesthesiologists and other physicians showed a 100% success rate of traditional scalpel approaches³⁸. However, the overall rate of surgical airways was lower than previously reported. These studies, though limited by their observational approach and low denominators, add to a growing amount of evidence that surgical cricothyrotomies may have a higher success rate when compared with cannula techniques^{39,40}.

Advocates of cannula techniques note that the needle-wire movements are more familiar to anesthesiologists, but cannula techniques may be vulnerable to a high rate of complications in the CICO scenario. The NAP4 analysis found that 90% of cricothyrotomies (26 out of 29) were first attempted with cannula techniques, potentially reflecting provider preference. However, 58% of cannula cricothyrotomies (15 out of 26) performed by anesthesiologists failed and required rescue¹. The cannula may provide a conduit for oxygenation via jet ventilation to temporize airway management, but barotrauma remains a significant risk in the obstructed airway. A systematic review of emergency transtracheal jet ventilation, including 90 out of 132 CICO scenarios (68%), showed higher rates of device failure, barotrauma, and overall complications in CICO emergency scenarios when compared with non-CICO emergencies⁴¹.

Different scalpel techniques

There is a lack of randomized data supporting any scalpel technique³⁶. Many techniques have similar steps (neck extension, identification of the cricothyroid membrane, incision through the skin and cricothyroid membrane, and insertion of a cuffed tracheal tube) with variations including the number of incisions and methods to keep the incision open. The 2015 Difficult Airway Society guidelines recommend a scalpel-bougie technique for tracheal cannulation with a gum-elastic bougie to keep the incision open^{16,36}. Other techniques, including the “4-step technique”, include devices such as tracheal hooks that may not be immediately available⁴². Training models in humans and animal cadaveric models have not favored one technique over another^{42–44}.

Given the challenges in surgical airway training, recent studies have investigated which techniques have higher success in both inexperienced and experienced learners. A recent study in surgical airway-naïve trainees showed a higher rate of success when using scalpel techniques versus cannula techniques on cadavers⁴⁵. Recent studies of needle versus scalpel techniques in animal cadaveric models show that senior anesthesiologists also achieve a higher rate of success with surgical scalpel techniques⁴⁶. However, more research is required before any definitive conclusions can be made about ease of learning with the different techniques.

Use of ultrasound

One of the more recent innovations in airway management is ultrasonography for the identification of landmarks for the surgical airway. The thought is that airway structures can be identified before induction of anesthesia so that emergency surgical airway management is performed with sound knowledge of the underlying airway anatomy. The growing number of obese patients significantly increases the likelihood of a difficult-to-assess surface anatomy. In addition, patients with previous head-and-neck trauma, surgery, or radiation (or a combination of these) may have difficult or impossible anatomy to identify by surface structure.

Patient selection for ultrasound identification of airway structures impacts its utility. A study of emergency medicine physicians and senior residents showed no significant difference in the ability to identify the cricothyroid membrane with palpation or ultrasound in healthy volunteers⁴⁷. In contrast, a study of women in labor showed that only 39% of anesthesiologists correctly identified the cricothyroid membrane in obese patients compared with 71% of non-obese patients⁴⁸. A study of overweight volunteers also showed a significant increase in ability to identify the cricothyroid membrane with ultrasound versus palpation (100% to 46%, respectively)⁴⁹. A study in human cadavers showed that naïve trainees were over five times more successful in cannulating the trachea with ultrasound guidance than palpation in difficult or impossible anatomy⁵⁰. So it appears that ultrasound is a useful tool for the identification of surgical airway structures in the obese patient population or others with difficult-to-identify surface airway structures.

Given the low incidence of CICO airways, most providers should be considered novice regardless of years in practice. Despite the

known benefits of both cognitive and procedural skills training in the CICO scenario, there are few standardized training programs. The current US Accreditation Council for Graduate Medical Education (ACGME) anesthesiology requirements do not require programs to train graduates in the surgical airway⁵¹. The ANZCA provides standards for CICO training sessions but does not clarify which algorithm is preferred⁵². In addition, such training programs are not often used by anesthesiologists or others who manage the airway, regardless of practice settings.

Future directions

As laryngoscopy devices continue to evolve, the literature continues to investigate whether these new devices improve intubation success and patient outcomes. Currently, initial data suggest that new techniques such as VL are not universally superior to traditional DL. Variations in setting, patient characteristics, situational stress, and provider training likely contribute to overall device success and are confounders that cannot all be modified.

A recent focus on CICO events and surgical airway management has highlighted the importance of simple, rapid techniques with ongoing training throughout one's career. Ultrasonography provides a helpful adjunct for the identification of surgical airway access anatomy when surface landmarks are unclear and should be introduced in surgical airway training.

Regardless of the technique, quick decision-making, familiarity with equipment, proper training, and effective communication remain the most important factors in a successful airway intervention. Fortunately, these characteristics can be improved with educational interventions and simulation training. As intubation techniques, devices, and evidence grow, airway education must continue to do so as well. The greatest opportunity for airway success remains in a straightforward plan with familiar equipment and practiced techniques.

Competing interests

The authors declare that they have no competing interests.

Grant information

The author(s) declared that no grants were involved in supporting this work.

References

1. Cook TM, Woodall N, Harper J, *et al.*: **Major complications of airway management in the UK: results of the Fourth National Audit Project of the Royal College of Anaesthetists and the Difficult Airway Society. Part 2: intensive care and emergency departments.** *Br J Anaesth.* 2011; 106(5): 632–42.
[PubMed Abstract](#) | [Publisher Full Text](#)
2. Semler MW, Janz DR, Lentz RJ, *et al.*: **Randomized Trial of Apneic Oxygenation during Endotracheal Intubation of the Critically Ill.** *Am J Respir*

Crit Care Med. 2016; 193(3): 273–80.

[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | [F1000 Recommendation](#)

3. Heard A, Toner AJ, Evans JR, *et al.*: **Apneic Oxygenation During Prolonged Laryngoscopy in Obese Patients: A Randomized, Controlled Trial of Buccal RAE Tube Oxygen Administration.** *Anesth Analg.* 2016.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
4. Patel A, Nouraei SA: **Transnasal Humidified Rapid-Insufflation Ventilatory**

- Exchange (THRIVE): a physiological method of increasing apnoea time in patients with difficult airways. *Anaesthesia*. 2015; 70(3): 323–9.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | [F1000 Recommendation](#)
5. **F** Chanques G, Riboulet F, Molinari N, *et al.*: Comparison of three high flow oxygen therapy delivery devices: a clinical physiological cross-over study. *Minerva Anesthesiol*. 2013; 79(12): 1344–55.
[PubMed Abstract](#) | [F1000 Recommendation](#)
 6. **F** Mitterlechner T, Herff H, Hammel CW, *et al.*: A dual-use laryngoscope to facilitate apneic oxygenation. *J Emerg Med*. 2015; 48(1): 103–7.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 7. **F** Mir F, Patel A, Iqbal R, *et al.*: A randomised controlled trial comparing transnasal humidified rapid insufflation ventilatory exchange (THRIVE) pre-oxygenation with facemask pre-oxygenation in patients undergoing rapid sequence induction of anaesthesia. *Anaesthesia*. 2016.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 8. **F** Malik MA, Subramaniam R, Maharaj CH, *et al.*: Randomized controlled trial of the Pentax AWS, Glidescope, and Macintosh laryngoscopes in predicted difficult intubation. *Br J Anaesth*. 2009; 103(5): 761–8.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 9. **F** Jungbauer A, Schumann M, Brunkhorst V, *et al.*: Expected difficult tracheal intubation: a prospective comparison of direct laryngoscopy and video laryngoscopy in 200 patients. *Br J Anaesth*. 2009; 102(4): 546–50.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 10. Enomoto Y, Asai T, Arai T, *et al.*: Pentax-AWS, a new videolaryngoscope, is more effective than the Macintosh laryngoscope for tracheal intubation in patients with restricted neck movements: a randomized comparative study. *Br J Anaesth*. 2008; 100(4): 544–8.
[PubMed Abstract](#) | [Publisher Full Text](#)
 11. **F** Wayne MA, McDonnell M: Comparison of traditional versus video laryngoscopy in out-of-hospital tracheal intubation. *Prehosp Emerg Care*. 2010; 14(2): 278–82.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 12. **F** Boehringer B, Choate M, Hurwitz S, *et al.*: Impact of Video Laryngoscopy on Advanced Airway Management by Critical Care Transport Paramedics and Nurses Using the CMAC Pocket Monitor. *Biomed Res Int*. 2015; 2015: 821302.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | [F1000 Recommendation](#)
 13. **F** Rhode MG, Vandborg MP, Bladt V, *et al.*: Video laryngoscopy in pre-hospital critical care - a quality improvement study. *Scand J Trauma Resusc Emerg Med*. 2016; 24: 84.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | [F1000 Recommendation](#)
 14. Trimmel H, Kreutziger J, Fertsak G, *et al.*: Use of the Airtraq laryngoscope for emergency intubation in the prehospital setting: a randomized control trial. *Crit Care Med*. 2011; 39(3): 489–93.
[PubMed Abstract](#) | [Publisher Full Text](#)
 15. **F** Trimmel H, Kreutziger J, Fitzka R, *et al.*: Use of the GlideScope Ranger Video Laryngoscope for Emergency Intubation in the Prehospital Setting: A Randomized Control Trial. *Crit Care Med*. 2016; 44(7): e470–6.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 16. **F** Hossfeld B, Frey K, Doerges V, *et al.*: Improvement in glottic visualisation by using the C-MAC PM video laryngoscope as a first-line device for out-of-hospital emergency tracheal intubation: An observational study. *Eur J Anaesthesiol*. 2015; 32(6): 425–31.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 17. **F** Naito H, Guyette FX, Martin-Gill C, *et al.*: Video Laryngoscopic Techniques Associated with Intubation Success in a Helicopter Emergency Medical Service System. *Prehosp Emerg Care*. 2016; 20(3): 333–42.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 18. **F** Sakles JC, Mosier JM, Patanwala AE, *et al.*: The C-MAC® video laryngoscope is superior to the direct laryngoscope for the rescue of failed first-attempt intubations in the emergency department. *J Emerg Med*. 2015; 48(3): 280–6.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 19. **F** Sakles JC, Patanwala AE, Mosier JM, *et al.*: Comparison of video laryngoscopy to direct laryngoscopy for intubation of patients with difficult airway characteristics in the emergency department. *Intern Emerg Med*. 2014; 9(1): 93–8.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 20. **F** Driver BE, Prekker ME, Moore JC, *et al.*: Direct Versus Video Laryngoscopy Using the C-MAC for Tracheal Intubation in the Emergency Department, a Randomized Controlled Trial. *Acad Emerg Med*. 2016; 23(4): 433–9.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 21. **F** Sulser S, Uebmann D, Schlaepfer M, *et al.*: C-MAC videolaryngoscope compared with direct laryngoscopy for rapid sequence intubation in an emergency department: A randomised clinical trial. *Eur J Anaesthesiol*. 2016; 33(12): 943–8.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 22. **F** Hypes CD, Stolz U, Sakles JC, *et al.*: Video Laryngoscopy Improves Odds of First-Attempt Success at Intubation in the Intensive Care Unit. A Propensity-matched Analysis. *Ann Am Thorac Soc*. 2016; 13(3): 382–90.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 23. **F** Mosier JM, Whitmore SP, Bloom JW, *et al.*: Video laryngoscopy improves intubation success and reduces esophageal intubations compared to direct laryngoscopy in the medical intensive care unit. *Crit Care*. 2013; 17(5): R237.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | [F1000 Recommendation](#)
 24. **F** Silverberg MJ, Li N, Acquah SO, *et al.*: Comparison of video laryngoscopy versus direct laryngoscopy during urgent endotracheal intubation: a randomized controlled trial. *Crit Care Med*. 2015; 43(3): 636–41.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 25. Griesdale DE, Chau A, Isac G, *et al.*: Video-laryngoscopy versus direct laryngoscopy in critically ill patients: a pilot randomized trial. *Can J Anaesth*. 2012; 59(11): 1032–9.
[PubMed Abstract](#) | [Publisher Full Text](#)
 26. **F** Janz DR, Semler MW, Lentz RJ, *et al.*: Randomized Trial of Video Laryngoscopy for Endotracheal Intubation of Critically Ill Adults. *Crit Care Med*. 2016; 44(11): 1980–7.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#) | [F1000 Recommendation](#)
 27. Lascarrou JB, Boissrame-Helms J, Bailly A, *et al.*: Video Laryngoscopy vs Direct Laryngoscopy on Successful First-Pass Orotracheal Intubation Among ICU Patients: A Randomized Clinical Trial. *JAMA*. 2017; 317(5): 483–493.
[PubMed Abstract](#) | [Publisher Full Text](#)
 28. **F** Joshi R, Hypes CD, Greenberg J, *et al.*: Difficult Airway Characteristics Associated with First Attempt Failure at Intubation Using Video Laryngoscopy in the Intensive Care Unit. *Ann Am Thorac Soc*. 2016.
[PubMed Abstract](#) | [F1000 Recommendation](#)
 29. **F** Kim JW, Park SO, Lee KR, *et al.*: Video laryngoscopy vs. direct laryngoscopy: Which should be chosen for endotracheal intubation during cardiopulmonary resuscitation? A prospective randomized controlled study of experienced intubators. *Resuscitation*. 2016; 105: 196–202.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 30. **F** Park SO, Kim JW, Na JH, *et al.*: Video laryngoscopy improves the first-attempt success in endotracheal intubation during cardiopulmonary resuscitation among novice physicians. *Resuscitation*. 2015; 89: 188–94.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 31. **F** Rosenstock CV, Thøgersen B, Afshari A, *et al.*: Awake fiberoptic or awake video laryngoscopic tracheal intubation in patients with anticipated difficult airway management: a randomized clinical trial. *Anesthesiology*. 2012; 116(6): 1210–6.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 32. **F** Kramer A, Müller D, Pförtner R, *et al.*: Fiberoptic vs videolaryngoscopic (C-MAC® D-BLADE) nasal awake intubation under local anaesthesia. *Anaesthesia*. 2015; 70(4): 400–6.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 33. Silvertown NA, Youngquist ST, Mallin MP, *et al.*: GlideScope versus flexible fiber optic for awake upright laryngoscopy. *Ann Emerg Med*. 2012; 59(3): 159–64.
[PubMed Abstract](#) | [Publisher Full Text](#)
 34. **F** Drenguis AS, Carlson JN: GlideScope vs. C-MAC for Awake Upright Laryngoscopy. *J Emerg Med*. 2015; 49(3): 361–8.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 35. **F** Abdellatif AA, Ali MA: GlideScope videolaryngoscope versus flexible fiberoptic bronchoscope for awake intubation of morbidly obese patient with predicted difficult intubation. *Middle East J Anaesthesiol*. 2014; 22(4): 385–92.
[PubMed Abstract](#) | [F1000 Recommendation](#)
 36. Frerk C, Mitchell VS, McNarry AF, *et al.*: Difficult Airway Society 2015 guidelines for management of unanticipated difficult intubation in adults. *Br J Anaesth*. 2015; 115(6): 827–48.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
 37. Baker PA, Flanagan BT, Greenland KB, *et al.*: Equipment to manage a difficult airway during anaesthesia. *Anaesth Intensive Care*. 2011; 39(1): 16–34.
[PubMed Abstract](#)
 38. Lockey D, Crewdson K, Weaver A, *et al.*: Observational study of the success rates of intubation and failed intubation airway rescue techniques in 7256 attempted intubations of trauma patients by pre-hospital physicians. *Br J Anaesth*. 2014; 113(2): 220–5.
[PubMed Abstract](#) | [Publisher Full Text](#)
 39. Hubble MW, Wilfong DA, Brown LH, *et al.*: A meta-analysis of prehospital airway control techniques part II: alternative airway devices and cricothyrotomy success rates. *Prehosp Emerg Care*. 2010; 14(4): 515–30.
[PubMed Abstract](#) | [Publisher Full Text](#)
 40. Mabry RL: An analysis of battlefield cricothyrotomy in Iraq and Afghanistan. *J Spec Oper Med*. 2012; 12(1): 17–23.
[PubMed Abstract](#)
 41. Duggan LV, Ballantyne Scott B, Law JA, *et al.*: Transtracheal jet ventilation in the 'can't intubate can't oxygenate' emergency: a systematic review. *Br J Anaesth*. 2016; 117(Suppl 1): i28–i38.
[PubMed Abstract](#) | [Publisher Full Text](#)
 42. Holmes JF, Panacek EA, Sakles JC, *et al.*: Comparison of 2 cricothyrotomy

- techniques: standard method versus rapid 4-step technique. *Ann Emerg Med.* 1998; **32**(4): 442–6.
[PubMed Abstract](#) | [Publisher Full Text](#)
43. Schaumann N, Lorenz V, Schellongowski P, *et al.*: **Evaluation of Seldinger technique emergency cricothyroidotomy versus standard surgical cricothyroidotomy in 200 cadavers.** *Anesthesiology.* 2005; **102**(1): 7–11.
[PubMed Abstract](#) | [Publisher Full Text](#)
 44. Hill C, Reardon R, Joing S, *et al.*: **Cricothyrotomy technique using gum elastic bougie is faster than standard technique: a study of emergency medicine residents and medical students in an animal lab.** *Acad Emerg Med.* 2010; **17**(6): 666–9.
[PubMed Abstract](#) | [Publisher Full Text](#)
 45. Heymans F, Feigl G, Graber S, *et al.*: **Emergency Cricothyrotomy Performed by Surgical Airway-naïve Medical Personnel: A Randomized Crossover Study in Cadavers Comparing Three Commonly Used Techniques.** *Anesthesiology.* 2016; **125**(2): 295–303.
[PubMed Abstract](#) | [Publisher Full Text](#)
 46. Chrisman L, King W, Wimble K, *et al.*: **Surgicric 2: A comparative bench study with two established emergency cricothyroidotomy techniques in a porcine model.** *Br J Anaesth.* 2016; **117**(2): 236–42.
[PubMed Abstract](#) | [Publisher Full Text](#)
 47. Yildiz G, Goksu E, Senfer A, *et al.*: **Comparison of ultrasonography and surface landmarks in detecting the localization for cricothyroidotomy.** *Am J Emerg Med.* 2016; **34**(2): 254–6.
[PubMed Abstract](#) | [Publisher Full Text](#)
 48. You-Ten KE, Desai D, Postonogova T, *et al.*: **Accuracy of conventional digital palpation and ultrasound of the cricothyroid membrane in obese women in labour.** *Anaesthesia.* 2015; **70**(11): 1230–4.
[PubMed Abstract](#) | [Publisher Full Text](#)
 49. Barbe N, Martin P, Pascal J, *et al.*: **[Locating the cricothyroid membrane in learning phase: value of ultrasonography?].** *Ann Fr Anesth Reanim.* 2014; **33**(3): 163–6.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 50. Siddiqui N, Arzola C, Friedman Z, *et al.*: **Ultrasound Improves Cricothyrotomy Success in Cadavers with Poorly Defined Neck Anatomy: A Randomized Control Trial.** *Anesthesiology.* 2015; **123**(5): 1033–41.
[PubMed Abstract](#) | [Publisher Full Text](#) | [F1000 Recommendation](#)
 51. Education ACfGM: **ACGME Program Requirements for Graduate Medical Education in Anesthesiology.** 2016; [cited 2016 November 20, 2016].
[Reference Source](#)
 52. Anaesthetists AaNZCo: **Appendix 12: Standards for Can't Intubate Can't Oxygenate (CICO) education sessions.** 2013; [cited 2016 November 20, 2016].
[Reference Source](#)

Open Peer Review

Current Peer Review Status:

Editorial Note on the Review Process

Faculty Reviews are review articles written by the prestigious Members of **Faculty Opinions**. The articles are commissioned and peer reviewed before publication to ensure that the final, published version is comprehensive and accessible. The reviewers who approved the final version are listed with their names and affiliations.

The reviewers who approved this article are:

Version 1

1. **Narasimhan Jagannathan**

Feinberg School of Medicine, Northwestern University, Chicago, IL, USA

Competing Interests: No competing interests were disclosed.

2. **Jarrod Mosier**

University of Arizona, Tuscon, AZ, USA

Competing Interests: No competing interests were disclosed.

The benefits of publishing with F1000Research:

- Your article is published within days, with no editorial bias
- You can publish traditional articles, null/negative results, case reports, data notes and more
- The peer review process is transparent and collaborative
- Your article is indexed in PubMed after passing peer review
- Dedicated customer support at every stage

For pre-submission enquiries, contact research@f1000.com

F1000Research