RESEARCH ARTICLE

A pilot study comparing pattern of damage sustained among instruments from different surgical units in a tertiary care centre in Nepal – reappraising the role of instrument reprocessing in retaining their value [version 1; peer review: 2 approved]

Sunil Munakomi ¹, Roshan Shah¹, Sangam Shrestha ²

¹Department of Neurosurgery, Nobel Hospital, Biratnagar, 0977, Nepal

²Department of Pediatrics, Koshi Zonal Hospital, Biratnagar, 0977, Nepal

V1 First published: 23 Jan 2018, 7:102
<https://doi.org/10.12688/f1000research.13699.1>
 Latest published: 23 Jan 2018, 7:102
<https://doi.org/10.12688/f1000research.13699.1>

Abstract

Background: The quality of instruments plays a pivotal role in governing safe operating room culture. The reprocessing system followed in the institution determines their durability thereby ensuring patient safety as well as minimizing health spending. Rigorous reprocessing in a centralized instrument reprocessing department by well trained staff following formulated guidelines helps to achieve the target of “safe surgery saves lives” as formulated by the World Health Organization.

Methods: We sought to determine the patterns of wear and tear sustained among sets of surgical equipment from two surgical units that had been sent to the repair department within a year of their purchase. Analysis of similar changes in the joints of the instrument, as well as pattern of fractures sustained was performed.

Results: All patterns of wear and tear were common in both the general surgical arm and neurosurgical counterpart, with the exception of fractures and mal-alignments. Similar study was performed examining changes in the joints. Stains were the most commonly observed change pattern in both sets of instruments. Fractures were most frequent in the working ends in both sets of instruments.

Conclusion: There is an alarming incidence of wear and tear patterns in the instruments used in the surgical units, even within the first year of their use. This supports the strict implementation of reprocessing guidelines by well trained workers and their quality assessments via audit checks. The quality of the purchased instruments also plays a pivotal role.

Open Peer Review

Approval Status

	1	2
version 1 23 Jan 2018	 view	 view

1. Vikal Chandra Shakya , Civil Service Hospital of Nepal, Kathmandu, Nepal
2. Samer S. Hoz , Neurosurgery Teaching Hospital, Baghdad, Iraq

Any reports and responses or comments on the article can be found at the end of the article.

Keywords

surgical, instruments, damage, pattern, reprocessing

Corresponding author: Sunil Munakomi (sunilmunakomi@gmail.com)

Author roles: **Munakomi S:** Conceptualization, Formal Analysis, Investigation, Methodology, Supervision, Validation, Writing – Original Draft Preparation, Writing – Review & Editing; **Shah R:** Formal Analysis, Methodology, Visualization; **Shrestha S:** Conceptualization, Formal Analysis, Investigation, Methodology, Writing – Review & Editing

Competing interests: No competing interests were disclosed.

Grant information: The author(s) declared that no grants were involved in supporting this work.

Copyright: © 2018 Munakomi S *et al.* This is an open access article distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited. Data associated with the article are available under the terms of the [Creative Commons Zero "No rights reserved" data waiver](#) (CC0 1.0 Public domain dedication).

How to cite this article: Munakomi S, Shah R and Shrestha S. **A pilot study comparing pattern of damage sustained among instruments from different surgical units in a tertiary care centre in Nepal – reappraising the role of instrument reprocessing in retaining their value [version 1; peer review: 2 approved]** F1000Research 2018, 7:102 <https://doi.org/10.12688/f1000research.13699.1>

First published: 23 Jan 2018, 7:102 <https://doi.org/10.12688/f1000research.13699.1>

Introduction

Surgical instruments are important assets to any surgical unit. The quality of this armamentarium links to the smooth running of operating theatres, as well as ensuring patient safety (see presentation on surface changes in surgical instruments [here](#)). The quality of their reprocessing minimizes the 'wear and tear' process, thereby ensuring their durability. It is paramount that formulated guidelines on the methods of reprocessing and sterilizing surgical instruments are implemented¹. This is more prudent in developing nations with limited resources, as it can help minimize their health costs. Simple steps of wiping out surgical soils (blood) from the instruments, use of demineralized water while rinsing, and proper storage contributes to better durability and improved patient safety. We, hereby, perform a pilot study focusing on the pattern of damage sustained in surgical instruments to determine the efficiency of instrument processing quality and instrument handling among different surgical units. This study aims to help reappraise the role of proper reprocessing and handling of surgical instruments, which is not well prioritized in the developing nations with limited resources. This would maintain the value of the instruments thereby increasing the durability and minimizing costs.

Methods

This study was carried out to determine the patterns of damage sustained among the instruments sent to the instrument maintenance unit of the Nobel Teaching Hospital, Nepal, from its different surgical units. Different sets of instruments were studied examining the pattern of changes seen with regards to spots, stains, pits, cracks, fractures, rust, mal-aligned parts, and tightening and loosening of the joints. Equal sets of instruments (47 each on comparable basis to the total of 47 neurosurgical instruments included in this study) sent to the institutional instrument maintenance unit from the departments of General Surgery and Neurosurgery, citing their repair or replacement within a year of their use, were analyzed. A comparative study was then conducted analyzing patterns in the 'wear and tear' process sustained among the instruments to provide some perspective on the quality of the reprocessing within the institution (see [AKI brochure](#) on instrument reprocessing techniques) ([Figure 1](#)–[Figure 3](#)).

The role of cleaning, handling and storing and the subsequent damages sustained among the instruments from the two surgical units was then studied. Further to this, a comparison of similar changes seen within the joints of the instruments from these departments was performed. The joint was chosen because of the high friction during usage, and the propensity for retained surgical soils within them, which predisposes to corrosion, stains, cracks and fractures. This was considered the signature marker of the quality of cleaning and reprocessing. Lastly, differences in the location of the fractures on the instruments was also documented ([Figure 4](#)).

This study was cleared by the Institutional Review Committee of Nobel Teaching Hospital, Biratnagar, Nepal.

Figure 1. Images showing patterns of common changes seen within the instruments.

Figure 2. Images showing pits, cracks and corrosion affecting the instruments.

Figure 3. Images showing damages sustained within the instruments under magnification.

Figure 4. Images showing different sites of fractures sustained within the instruments.

Results

Patterns of damage in the general surgical instruments

Stains was the most common observation, seen in 97.87% of the instruments, followed by loosening in 82.97%, rust in 27.65%, pits in 25.5% and mal-alignment in 19% of the instruments ([Figure 5](#)).

When assessing the joints, loosening was seen in 82.97%, stains in 80.85%, rust in 29.78% and tightening in 14.89% of the instruments (Figure 6).

Fractures were seen in 17.02% of the instruments, mostly found at the tip (75%) followed by involvement of the shaft and handle (12.5% each).

Pattern of damage in the neurosurgical instruments

Stains were the most common observation, found in 38.29% of instruments, followed by loosening in 31.91%, mal-alignment in 29.78% and discoloration in 23.4% (Figure 7).

When assessing the joints, loosening and staining were seen in 29.78% each, whereas rust was found in only 8.5% (Figure 8).

Fractures were seen in 27.65%, mostly at the tip (46.1%), followed by involvement of the joint in 23% and shaft and handle in 15.38% each.

Comparison of the damage pattern among the two units

Comparing the patterns of damage among the instruments from the two different surgical units, all patterns were higher in the general surgical arm, except for the incidence of fractures (27.65% vs 17.02%) and mal-alignment (29.78% vs 19%) (Figure 9).

Figure 5. Pie chart representing different patterns of wear and tear sustained in the instruments from the department of general surgery.

Figure 6. Pie chart showing changes seen within the joints of the instruments from the department of General surgery.

Examining the damage in the joint areas, more damage was seen in the instruments from the general arm with the exception of fractures sustained (6.3% vs 0%) (Figure 10).

Figure 7. Pie chart representing patterns of changes and damages seen within the instruments from the department of neurosurgery.

Figure 8. Pie chart showing changes within the joints of the instruments from Neurosurgery.

Figure 9. Chart showing correlation between patterns of damage sustained between the instruments from general surgery (blue) and neurosurgery (red). Arrows indicate where the incidence of a specific change is higher in neurosurgery compared to general surgery.

Figure 10. Chart showing correlation between patterns of damage sustained within the joints of the instruments between general surgery (blue) and neurosurgery (red). Arrows indicate where the incidence of a specific change is higher in neurosurgery compared to general surgery.

The most common site of fractures in instruments from both units was in the tip (75% vs 46.1%).

Dataset 1. Tables listing all observed patterns of damage for each instrument surveyed in the study

<http://dx.doi.org/10.5256/f1000research.13699.d190910>

Discussion

Spaulding classified instruments into critical, semi-critical and non-critical depending upon the risk involved in transmitting infections through their usage². Recent studies have proven the presence of significant amount of stains, soils, structural damage as well as bio-films in ready to use instruments despite multiple stages of processing³. Despite the recent advancements made in the medical field, infection transmission following improper reprocessing is still a major concern^{4,5}. Proper processing of instruments therefore ensures better durability and improves patient safety. With this in mind, the reprocessing unit should precisely follow the manufacturer's instructions and the relevant institutional guidelines (see presentation on surface changes in surgical instruments [here](#)). It is therefore, prudent to establish, standardize and audit the guidelines on reprocessing of medical devices⁶. This ensures better safety for patients and health workers as well as enhancing the value retention of the instruments (see presentation on surface changes in surgical instruments [here](#)).

Recent guidelines recommend implementing compliance with set protocols in a Central Sterile Supply Department (CSSD), with trained staff, in a controlled environment. The whole process should be periodically counter checked and well audited⁶. All steps involved, such as disassembling, sorting, soaking, cleaning, rinsing, drying, reassembling, inspecting, lubricating, and wrapping, should be properly implemented⁶.

The Instrument Reprocessing Working Group (AKI) has recently formed a red brochure that details all damages that can occur during processing and sterilization, as well as the means to minimize them (see [AKI brochure](#) on instrument reprocessing techniques). Impurities in the rinsing water can lead to stains (silicates), pitting (chlorides), and blackening (ammonium ions). Improper care of the joints of the instruments can lead to corrosion, cracks and thereby facilitate early fractures. Inadequate drying can lead to rusting in the instruments. The use of demineralized water, proper cleaning, and avoiding dampness ensures better protection from spots, stains, rusts and pits, thereby increasing their durability and minimizing health costs (see [AKI brochure](#) on instrument reprocessing techniques). Proper handling, correct loading, ensuring material compatibility, timely lubrication (milking) and periodic checks for surface integrity further ensure their maintenance. They have further recommended a maximum waiting period of 6 hours between usage of instruments and start of cleaning, and as well as usage of demineralized water in the reprocessing cycle (see [AKI brochure](#) on instrument reprocessing techniques).

Our study showed that stains were the most common changes seen in the instruments. This suggests either improper wiping off of soils or the deposition of impurities in the rinsing processes. Low incidence of such changes in the neurosurgical unit indicates a better reprocessing attitude in the unit. However, better quality checks of the rinsing water for impurities, and better clearing of the surgical soils retains paramount value in reprocessing cycle. The high incidence of fractures and mal-alignment in the neurosurgical sets may be attributable to the frequent usage of fine micro-instruments with sharp working ends with fine joints and springs in the department. Better care of such fine and expensive equipment may also have resulted in them faring better in comparison to the general surgery counterpart. The high incidence of stains rusts and loosening seen among the instruments from the general surgery units may be attributable to their long usage, and also their sharing among different units before the procurement of new sets. Instruments in neurosurgery are used by limited surgeons and there is a propensity for faster replacement, owing to the damage to the fine working ends. Chronic changes such as rust, cracks and corrosions are observed less in the neurosurgical unit. One of the limiting factors of our study can be the quality of the instruments purchased in the departments. However, any surgical instrument is said to have shelf-life of at least 10 years if adequate care of the instruments is taken⁷.

The World Health Organization (WHO) has already implemented the notion of "Safe Surgery Saves Lives", and formulated strict guidelines governing them⁸. Safe operating room culture can prevent sentinel events that can prove to be devastating as well as help minimize malpractice crisis⁹. The 'cleanability' and configuration of instruments bears immense impact upon the patient safety^{10,11}. The Joint Commission found 74% of all "immediate threat to life declarations" were directly related to improperly sterilized instruments. They therefore formulated a checklist to minimize this (see [AKI brochure](#) on instrument reprocessing techniques). One study has shown major concerns regarding the effectiveness of sterilization being practiced in low and middle income countries¹². Despite the prevalence of such

poor practices in developing countries, proper assessment tools can eliminate these loop-holes¹³. A curriculum based on patient safety needs to be addressed and practiced for ensuring patient safety¹⁴. Such practices promote safe surgical practice^{15,16}. This also prevents the formation of bio-films on the surfaces of surgical instruments^{17,18}. There is utmost need for a paradigm shift in sterile processing departments, and the implementation of new approaches in processing^{19,20}. This requires acquisition of a thoroughly trained workforce for the process²¹. Paradoxically unqualified workers are more often placed in these roles, thereby jeopardizing the whole reprocessing cycle²². Furthermore, continuous and systematic quality improvement monitoring needs to be implemented²³.

Conclusion

Proper instrument handling and their reprocessing needs be of primary importance. Strict guidelines need to be formulated and strictly implemented in a centralized area in a dogmatic fashion. This is even more prudent in the context of resource limited setups. This culture promotes the practice of safe surgery and thereby maximizes patient's safety. Our study also highlighted the fact that simple steps in the reprocessing such as cleaning,

handling, wrapping and storage of surgical instruments can have significant impact on the overall durability of the instruments, which are the working hands of any surgeons. Care also needs to be given on purchasing quality instruments and also for allocating trained manpower for this critical reprocessing cycle. Such practice can minimize health investments thereby allowing improved allocation of limited health resources available in developing countries like ours. This would also help minimize the malpractice crisis that is slowly lurking in the health sector in the global front.

Data availability

Dataset 1: Tables listing all observed patterns of damage for each instrument surveyed in the study [10.5256/f1000research.13699.d190910](https://doi.org/10.5256/f1000research.13699.d190910)²⁴.

Competing interests

No competing interests were disclosed

Grant information

The author(s) declared that no grants were involved in supporting this work.

References

1. Canadian Standards Association: **CAN/CSA-Z314.0-13 Medical Device Reprocessing - General requirements**. Rexdale, Ont.: Canadian Standards Association; 2013.
[Reference Source](#)
2. Spaulding E, editor: **The role of chemical disinfection in the prevention of nosocomial infections**. International Conference on Nosocomial Infections; 1970 1971; Chicago, IL: American Hospital Association.
3. de Melo Costa D, de Oliveira Lopes LK, Tipple AF, *et al.*: **Evaluation of stainless steel surgical instruments subjected to multiple use/processing**. *Infection, Disease & Health*. (Ahead of print).
[Publisher Full Text](#)
4. Lipscomb IP, Sihota AK, Keevil CW: **Comparative Study of Surgical Instruments from Sterile-Service Departments for Presence of Residual Gram-Negative Endotoxin and Proteinaceous Deposits**. *J Clin Microbiol*. 2006; **44**(10): 3728–3733.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
5. Dancer SJ, Stewart M, Coulombe C, *et al.*: **Surgical site infections linked to contaminated surgical instruments**. *J Hosp Infect*. 2012; **81**(4): 231–8.
[PubMed Abstract](#) | [Publisher Full Text](#)
6. **Best Practices for Cleaning, Disinfection and Sterilization of Medical Equipment/Devices in All Health Care Settings**, 3rd edition. 2013.
[Reference Source](#)
7. Spry CC: **Care and Handling of Basic Surgical Instruments**. *AORN J*. 2007; **86**: S77–S81.
[Publisher Full Text](#)
8. World Health Organization: **Safe surgery saves lives**. 2013.
[Reference Source](#)
9. Makary MA, Sexton JB, Freischlag JA, *et al.*: **Patient Safety in Surgery**. *Ann Surg*. 2006; **243**(5): 628–635; discussion 632–5.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
10. Chobin N, Swanson K: **Putting Patient Safety First: The Sterile Processing Department and Healthcare Technology Management**. *Biomed Instrum Technol*. 2012; **46**(Suppl): 27–31.
[PubMed Abstract](#) | [Publisher Full Text](#)
11. Scheinberg N, Zhang B, Raschid L, *et al.*: **A Systematic Approach to Improve the Reprocessing of Surgical Instruments**. In: Duffy V, Lightner N. (eds) *Advances in Human Factors and Ergonomics in Healthcare*. Advances in Intelligent Systems and Computing, Springer, Cham, 2017; **482**: 275–286.
[Publisher Full Text](#)
12. O'Hara NN, Patel KR, Caldwell A, *et al.*: **Sterile reprocessing of surgical instruments in low- and middle-income countries: A multicenter pilot study**. *Am J Infect Control*. 2015; **43**(11): 1197–200.
[PubMed Abstract](#) | [Publisher Full Text](#)
13. Fast O, Fast C, Fast D, *et al.*: **Limited sterile processing capabilities for safe surgery in low-income and middle-income countries: experience in the Republic of Congo, Madagascar and Benin**. *BMJ Global Health*. 2017; **2**(Suppl 4): e000428.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
14. Kim FJ, da Silva RD, Gustafson D, *et al.*: **Current issues in patient safety in surgery: a review**. *Patient Saf Surg*. 2015; **9**: 26.
[PubMed Abstract](#) | [Publisher Full Text](#) | [Free Full Text](#)
15. Seavey R: **High-level disinfection, sterilization, and antisepsis: current issues in reprocessing medical and surgical instruments**. *Am J Infect Control*. 2013; **41**(5 Suppl): S111–S117.
[PubMed Abstract](#) | [Publisher Full Text](#)
16. Seavey RE: **Safe instrument reprocessing: the perioperative role**. *AORN J*. 2015; **101**(4): 482–5.
[PubMed Abstract](#) | [Publisher Full Text](#)
17. **Guideline for cleaning and care of surgical instruments**. In: *Guidelines for Perioperative Practice*. Denver, CO: AORN, Inc; 2015: 615–650.
18. Association for the Advancement of Medical Instrumentation: **ANSI/AAMI ST79: Comprehensive guide to steam sterilization and sterility assurance in health care facilities**. Arlington, VA: Association for the Advancement of Medical Instrumentation; 2013.
[Reference Source](#)
19. Swanson SC: **Shifting the sterile processing department paradigm: a mandate for change**. *AORN J*. 2008; **88**(2): 241–7.
[PubMed Abstract](#) | [Publisher Full Text](#)
20. Blyth PL, Shimkus J: **New tools need new approaches. The artistry of SPD (sterile processing department)**. *Mater Manag Health Care*. 2000; **9**(7): 24–5.
[PubMed Abstract](#)
21. Garner M, Collins A, Dunphy S, *et al.*: **Developing a central sterile surgical instrument technician program**. *AORN J*. 2004; **80**(2): 235–8, 241.
[PubMed Abstract](#) | [Publisher Full Text](#)
22. Tipple AF, de Souza TR, Bezerra AL, *et al.*: **[The worker with no training in nursing in a center of materials and sterilization: a challenge to the nurse]**. *Rev Esc Enferm USP*. 2005; **39**(2): 173–180.
[PubMed Abstract](#) | [Publisher Full Text](#)

23. Choo S, Papandria D, Goldstein SD, *et al.*: **Quality improvement activities for surgical services at district hospitals in developing countries and perceived barriers to quality improvement: findings from Ghana and the scientific literature.** *World J Surg.* 2013; 37(11): 2512–9.
[PubMed Abstract](#) | [Publisher Full Text](#)
24. Munakomi S, Shah R, Shrestha S: **Dataset 1 in: A pilot study comparing pattern of damage sustained among instruments from different surgical units in a tertiary care center in Nepal – reappraising the role of instrument reprocessing in retaining their value.** *F1000Research.* 2018.
[Data Source](#)

Open Peer Review

Current Peer Review Status:

Version 1

Reviewer Report 26 September 2018

<https://doi.org/10.5256/f1000research.14881.r38035>

© 2018 Hoz S. This is an open access peer review report distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Samer S. Hoz

Department of Neurosurgery, Neurosurgery Teaching Hospital, Baghdad, Iraq

The article represents a brief and concise review. The topic under focus "surgical instruments-care improvement" is an interesting topic and contributes valuable added knowledge to the field. It is written in an accepted language skill with consistent attention to the needs of readers in a clear and comprehensive organization.

The strength of the article is that it presents a clear summation of a critical daily surgical concerns. Its weakness may be that it does not offer adequate solutions rather than highlighting the different forms and patterns of damages.

In summary, the central concepts clearly defined within the article with adequate evidence, fits the journal field and contributes to advance the knowledge in the surgical field.

Is the work clearly and accurately presented and does it cite the current literature?

Yes

Is the study design appropriate and is the work technically sound?

Yes

Are sufficient details of methods and analysis provided to allow replication by others?

Yes

If applicable, is the statistical analysis and its interpretation appropriate?

Yes

Are all the source data underlying the results available to ensure full reproducibility?

Yes

Are the conclusions drawn adequately supported by the results?

Yes

Competing Interests: No competing interests were disclosed.

I confirm that I have read this submission and believe that I have an appropriate level of expertise to confirm that it is of an acceptable scientific standard.

Reviewer Report 06 February 2018

<https://doi.org/10.5256/f1000research.14881.r30137>

© 2018 Shakya V. This is an open access peer review report distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Vikal Chandra Shakya

Department of Surgery, Civil Service Hospital of Nepal, Kathmandu, Nepal

The article is well-written with comprehensive review of the literature; with good conclusions; with definitely a contribution to practice of the operating surgical units. Good evaluation of the results is another best aspect of this article. The conceptualization and analysis is commendable with well written English. This article would definitely support assessment of the the purchased instruments to be applied in the future.

Is the work clearly and accurately presented and does it cite the current literature?

Yes

Is the study design appropriate and is the work technically sound?

Yes

Are sufficient details of methods and analysis provided to allow replication by others?

Yes

If applicable, is the statistical analysis and its interpretation appropriate?

Yes

Are all the source data underlying the results available to ensure full reproducibility?

Yes

Are the conclusions drawn adequately supported by the results?

Yes

Competing Interests: No competing interests were disclosed.

Reviewer Expertise: General and gastrotintestinal surgery

I confirm that I have read this submission and believe that I have an appropriate level of expertise to confirm that it is of an acceptable scientific standard.

The benefits of publishing with F1000Research:

- Your article is published within days, with no editorial bias
- You can publish traditional articles, null/negative results, case reports, data notes and more
- The peer review process is transparent and collaborative
- Your article is indexed in PubMed after passing peer review
- Dedicated customer support at every stage

For pre-submission enquiries, contact research@f1000.com

F1000Research