METHOD ARTICLE

REVISED Identification of nitrite treated tuna fish meat via the determination of nitrous oxide by head space-gas chromatography/mass spectrometry [version 2; peer review: 2 approved]

Markus Niederer , Sandra Lang, Bernard Roux, Thomas Stebler, Christopher Hohl

State Laboratory Basel-City, Kannenfeldstr. 2, 4056, Basel, Switzerland

v2 First published: 22 May 2019, 8:711
<https://doi.org/10.12688/f1000research.19304.1>
Latest published: 12 Aug 2019, 8:711
<https://doi.org/10.12688/f1000research.19304.2>

Abstract

Tuna fish meat is an expensive and highly perishable sea food. Fresh meat has a bright red colour which soon turns into an unsightly brown during storage. To prolong the aspect of freshness, the red colour is stabilised or even enhanced e.g. with carbon monoxide or nitric oxide, the product of a nitrite / ascorbic acid treatment, which bind as a ligand to myoglobin. These procedures are illegal. Here we present a method for identifying tuna meat samples, which have undergone fraudulent wet salting with nitrite. The method uses headspace-gas chromatography/mass spectrometry (GC/MS) for the determination of nitrous oxide, which is formed as the final product of the two-step reduction nitrite (added agent) to nitric oxide (ligand) to nitrous oxide (target compound). Complex bound nitric oxide is set free with sulfuric acid, which also promotes the reduction to nitrous oxide. The method was validated using ^{15}N labelled nitrite as well as treated and untreated reference fish samples. A survey of 13 samples taken from the Swiss market in 2019 showed that 45 % of all samples were illegally treated with nitrite.

Keywords

tuna, red colour stability, food fraud, nitrite, nitrous oxide, food additives

This article is included in the **Agriculture, Food and Nutrition** gateway.

Open Peer Review

Approval Status

	1	2
version 2 (revision) 12 Aug 2019		
version 1 22 May 2019	view	view

1. **Ulrich Busch**, The Bavarian State Office for Health and Food Safety (LGL), Erlangen, Germany
2. **Maurus Biedermann**, Official Food Control Authority of the Canton of Zurich (Kantonales Labor), Zürich, Switzerland

Any reports and responses or comments on the article can be found at the end of the article.

Corresponding author: Markus Niederer (markus.niederer@bs.ch)

Author roles: **Niederer M:** Conceptualization, Data Curation, Formal Analysis, Methodology, Project Administration, Resources, Validation, Visualization, Writing – Original Draft Preparation, Writing – Review & Editing; **Lang S:** Data Curation, Investigation, Validation; **Roux B:** Data Curation, Investigation, Methodology, Validation; **Stebler T:** Conceptualization, Investigation, Methodology, Validation; **Hohl C:** Conceptualization, Project Administration, Supervision, Writing – Original Draft Preparation, Writing – Review & Editing

Competing interests: No competing interests were disclosed.

Grant information: The author(s) declared that no grants were involved in supporting this work.

Copyright: © 2019 Niederer M *et al.* This is an open access article distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

How to cite this article: Niederer M, Lang S, Roux B *et al.* **Identification of nitrite treated tuna fish meat via the determination of nitrous oxide by head space-gas chromatography/mass spectrometry [version 2; peer review: 2 approved]** F1000Research 2019, 8:711 <https://doi.org/10.12688/f1000research.19304.2>

First published: 22 May 2019, 8:711 <https://doi.org/10.12688/f1000research.19304.1>

REVISED Amendments from Version 1

Implementing reviewer comments we now indicate, that the method also allows for the simultaneous determination of CO and N₂O in the same GC run. We further specified the type of calibration used (one-point) and the limit of detection for N₂O.

See referee reports

Introduction

As Tuna fish species have been overfished on the one hand, and their meat being a sought-after but highly perishable food stuff on the other hand, the seafood industry has felt itself constrained to undertake efforts in increasing the shelf life of this commodity. Being a bright red in its fresh state, but turning into an unsightly brown during storage, attempts have also been made in extending or even enhancing the attractive aspect of fresh meat. The red colour of meat in its natural state is due to myoglobin in its reduced form (Mb-Fe^{II}-O₂). Being susceptible to autoxidation, the resulting metmyoglobin (Mb-Fe^{III}-O₂) is brown in colour. One way of stabilising the red colour is by gassing with carbon monoxide (CO), which binds as a ligand to myoglobin (Mb-Fe^{II}-CO). A second way consists in applying the anion nitrite (NO₂⁻) in combination with the reducing agent ascorbic acid to form nitric oxide (NO), which forms a complex with both myoglobin (Mb-Fe^{II}-NO) and metmyoglobin (Mb-Fe^{III}-NO)¹. Whereas the colour hue of the CO complex is described as cherry red, that of the NO-complexes are described as pink red. Neither procedure is authorised for fresh fish². However, with these possibilities at hand, abuse for simulating the fresh aspect of tuna fish meat is a tempting option^{3,4}, all the more so regarding the currently insufficient tools for fraud detection: Whereas, for CO, analytical methods are at hand to support law enforcement^{5,6}, attempts to detect nitrite treated meat, either by searching for nitrite residues or traces of nitrosamines, have not been reported as being successful yet⁷. Here we report on our method for the detection of nitrite

treatment (brining), which originates from the method for CO determination and determines nitrous oxide (N₂O) as the reduction product of nitric oxide (NO), stemming from the complex Mb-Fe^{II}-NO (Figure 1). This approach allows the determination of CO and N₂O even within the same GC/MS run.

Methods

Chemicals

Sulphuric acid (95–97%, cat. no. 1.00731), 1-octanol (puriss., cat. no. 1.00991), sodium chloride (NaCl, p.a., cat. no. 1.06404.1000), sodium nitrite (NaNO₂, p.a., cat. no. 1.06549.0100), sodium nitrite-N15 (Na¹⁵NO₂, 99%, cat. no. 490814-1G), ascorbic acid (p.a., cat. no. 127) were all from Merck (Buchs, Switzerland) and sodium ascorbate (99%, cat. no. 352681000) from Acros (New Jersey, U.S.A.). Certified NO- gas (100 ppmv N₂, 5 L, 150 bar, cat. no. 1878) was purchased from Carbogas (Basel, Switzerland) and N₂O- gas (100 ppmv N₂, 2 L, 150 bar, cat. no. 7092325) was from Messer AG (Lenzburg, Switzerland).

Samples and sample preparation

Tuna samples (frozen loins, tuna saku, loose and pre-prepared packages) were taken from retailers and importers in Basel and stored at – 18°C until sample preparation. An aliquot of 15 g was weighed into a 70 mL centrifuge tube and 15 g of crushed ice as well as 15 mL of water were added. The mixture was homogenised at 7,000–9,000 rpm with a polytron homogeniser (Kinematica AG, Switzerland) and centrifuged for 5 min at 3,500 rpm. An aliquot of the resulting supernatant (5 mL) was transferred into a 20 mL headspace vial. After adding 5 µL of n-octanol and 2 mL of sulphuric acid (20 % in deionised water) the vial was immediately sealed and gently shaken by hand for 30 s. Certified NO- gas 100 ppmv (122.6 ng/mL) and N₂O- gas 100 ppmv in N₂ (180.6 ng/mL) served as standards for a one-point calibration: A headspace vial was filled with 5 mL of deionised water and 2 mL of sulphuric acid (20 %), the cap only loosely fitted and the reference gas introduced with a needle through the septum. The cap was then sealed tight.

Figure 1. Postulated reaction schemes: brining with NO₂⁻; formation of N₂O as a stable reduction product of NO. In bold: verification experiment with labelled Na¹⁵NO₂ and resulting ¹⁵N₂O.

Headspace – gas chromatography/mass spectrometry (GC/MS)

The sealed vials were placed into the incubation unit of the CombiPAL autosampler (CTC Analytics, Zwingen, Switzerland) for 60 min at 25°C and analysed with a Trace 1310 GC / ISQ quadrupole system (Thermo Scientific, U.S.A.) under the following conditions: 500 µL of headspace was injected at 70°C (split 1:20) onto a Porous Layer Open Tubular (PLOT) column (30 m × 0.32 mm inner diameter, 12 µm film thickness, HP-PLOT Molsieve, Agilent Technologies, cat. no. 19091P-MS4), Palo Alto, CA, U.S.A.). Helium (99.999%, Carbagas, Basel, Switzerland, cat. no. P6201RG) was used as carrier gas at a constant flow rate of 1 mL/min. The oven temperature was programmed from 70°C (5 min) to 220°C (3 min) at 30°C/min. The quadrupole detector was set in the electronic ionisation mode (full scan 10–100 amu) with selected-ion monitoring (70 eV, ion source = 170°C, delay = 2.5 min) with a transfer-line to the GC (1 m fused silica deactivated, 0.25 mm i.d.) at 200°C. Specific masses for the detection were m/z 30 for NO (m/z 31 for ^{15}NO) at a retention time of about 7.9 minutes and m/z 44 for N_2O (m/z 46 for $^{15}\text{N}_2\text{O}$) at 11.4 minutes.

Brining experiments

For brining experiments, 15 g of nine untreated fish samples (certified as 100 % tuna from the Maldives, the Philippines, China, Vietnam) were each cut into pieces of about 1 cm and treated with 20 mL of brine solution (10% NaCl and 0.4% NaNO_2 in deionised water)⁸ at room temperature for 24 hours. After removing the solution, the now reddish looking tuna samples along with eight untreated control samples were each bathed twice in 15 mL of antioxidant solution (3.0 % ascorbic acid and sodium ascorbate in deionised water) for one hour at room temperature and subsequently washed twice with antioxidant solution. Thereafter, samples were prepared as described above and analysed by headspace-GC/MS.

Verification experiment: The whole procedure was also done with three tuna samples using $\text{Na}^{15}\text{NO}_2$ for the brine solution instead.

Results and discussion

Method development

In contrast to carbon monoxide, where the agent used for treatment, the ligand responsible for colour stabilisation and the target compound for sample screening is one and the same molecule, nitrite treatment identification has to consider the two-step reduction from nitrite to nitric oxide to nitrous oxide. This required changing the original GC temperature programme of the CO method: Instead of an isothermic run at 40°C, oven temperature was ramped up from 70°C to 220°C.

Both carbon monoxide and nitric oxide are bound in the sample as a ligand. Their release is enhanced by adding sulfuric acid. Being a free radical, however, nitric oxide is far more reactive than carbon monoxide: Instead of a sharp symmetrical peak, chromatograms show nitric oxide as a broad hump, barely distinguishable from baseline noise. The direct approach therefore is not feasible. In an acidic environment, however, two NO molecules, upon reduction, react to nitrous oxide (N_2O), which shows up as a perfect peak in the chromatogram (Figure 2), allowing for the indirect detection of NO. For the purpose of identifying nitrite treatment, a semi quantitative assessment of N_2O proved to be sufficient, as any clear signal is alone due to a former nitrite treatment.

Brining experiments

Verification of this mechanism was performed using isotope labelled nitrite ($\text{Na}^{15}\text{NO}_2$) for brining (Figure 1). During subsequent analysis, both the ^{15}NO hump and $^{15}\text{N}_2\text{O}$ peak were detected (Figure 2), which is consistent with the postulated reaction pathway. In addition, brining experiments with nine tuna samples using unlabelled NaNO_2 showed sharp N_2O signals in the range of 1000 to 12000 µg/kg (Figure 3, underlying

Figure 2. Chromatograms and mass spectra of native $^{14}\text{N}_2\text{O}$ as well as labelled $^{15}\text{N}_2\text{O}$ and ^{15}NO released from nitrite treated samples.

Figure 3. Box-and-Whisker-Plots of brining experiments. $^{15}\text{N}_2\text{O}$ labelled samples (x).

data⁹). In contrast, untreated reference samples, even though they were exposed to antioxidants, were free of nitrous oxide (limit of detection about 30 µg/kg). Therefore, naturally occurring relevant concentrations of N_2O can be excluded.

Validation with reference samples

The method was validated with seven reference tuna samples of a trustworthy source from 2017 which were stored at -18°C for two years in our lab (see underlying data¹⁰). Five of them were declared as untreated and consists of 100% tuna. The remaining two samples were declared as treated with rosemary extract (E392) and/or nitrite and labelled with additives such as salt and antioxidants (E300, E301, E331). The untreated samples were free of N_2O (< 30 µg/kg) whereas the treated ones showed stable N_2O signals of 260 and 750 µg/kg. These results not only confirm our findings from the brining experiments, but are also consistent with the corresponding documents of the reference tuna samples.

Market survey

The method was then used on a routine basis in a market survey of 13 samples of raw tuna collected in Basel in 2019 (see underlying data¹¹). Seven samples (54%) were negative. Six samples (46%) were tested positive with N_2O peaks corresponding to levels of between 90 and 1130 µg/kg. As a consequence, these six samples, all originating from Vietnam, were objected according to law.

Conclusion

With our experiments using labelled $\text{Na}^{15}\text{NO}_2$ and the comparison of our results with the corresponding documents of the reference tuna samples, we could demonstrate that our

method using nitrous oxide as a target compound is suitable for identifying nitrite manipulated tuna. Its use in a market survey also showed, that the method is fit for purpose. In addition, the method was even capable of identifying nitrite treatment in samples stored at -18°C for two years before analysis.

Data availability

As a law enforcement body, we are bound to official secrecy. Considering this restriction, names of products and importers were anonymised.

Figshare: Fig03 brining experiment rawdata.csv. <https://doi.org/10.6084/m9.figshare.8142992.v1>⁹

This project contains the following underlying data:

- Fig03 brining experiment rawdata.csv (Raw data and descriptive statistic data (univariate diagram) of "figure 03" performed with the Add-In XLSTAT 2009.1.02 is provided as Excel-file (CSV). The data include file name, sample name, type of treatment, area, calculated N_2O amounts and statistical values)

Figshare: Validation with Reference samples 2017. <https://doi.org/10.6084/m9.figshare.8143016.v1>¹⁰

This project contains the following underlying data:

- Validation with Reference samples 2017.csv (Raw data of 7 reference samples of 2017 of the "validation section" are provided as Excel-file (CSV). The data include file name, sample name, certified type of treatment and calculated N_2O amounts (single and average values))

Figshare: Market survey 2019 rawdata. <https://doi.org/10.6084/m9.figshare.8143031.v1>¹¹

This project contains the following underlying data:

- Market survey 2019 rawdata.csv (Raw data and descriptive statistic data of the "market survey section" performed with the Add-In XLSTAT 2009.1.02 are provided as Excel-file (CSV). The data include file name, sample name, area, calculated N2O amounts, test result and statistical values)

Data are available under the terms of the [Creative Commons Zero "No rights reserved" data waiver](#) (CC0 1.0 Public domain dedication).

Grant information

The author(s) declared that no grants were involved in supporting this work.

References

1. Faustman C: **Myoglobin Chemistry and Modifications that Influence (Color and) Color Stability**. In: *American Meat Science Association, 67th Annual Reciprocal Meat Conference*, 2014.
[Reference Source](#)
2. **Verordnung des EDI über die zulässigen Zusatzstoffe in Lebensmitteln**. (817.022.31), Bern, 2017.
[Reference Source](#)
3. **Operation OPSON VII (2018)**. Commission DG Health and Food Safety, 2018.
[Reference Source](#)
4. **Bericht - OPSON VII: Wurde der Thunfisch «schöngefärbt»? Bundesamt für Lebensmittelsicherheit und Veterinärwesen (BLV)**, Bern, 2018.
[Reference Source](#)
5. Anderson RC, Wu WH: **Analysis of carbon monoxide in commercially treated tuna (*Thunnus spp.*) and mahi-mahi (*Coryphaena hippurus*) by gas chromatography/mass spectrometry**. *J Agric Food Chem*. 2005; **53**(18): 7019–23.
[PubMed Abstract](#) | [Publisher Full Text](#)
6. Frey T, Roux B, Eymann W: **Detection of Carbon Monoxide-Treated Tuna by Headspace-Gas Chromatography/Mass Spectrometry**. *Chimia*. 2006; **60**: 241.
[Publisher Full Text](#)
7. Niedersächsischen Landesamt für Verbraucherschutz und Lebensmittelsicherheit (LAVES, IFF) Cuxhaven: **Ein Thunfisch sieht rot - Verbrauchertäuschung durch farbstabilisierten Thunfisch**. 2018.
[Reference Source](#)
8. Aydin K, Aydin O: **Effect of Ascorbic Acid Utilization on Cold Smoked Fish Quality (*Oncorhynchus mykiss*) during Process and Storage**. *Food Sci Technol Res*. 2013; **19**(5): 823–831.
[Publisher Full Text](#)
9. Niederer M: **Fig03 brining experiment rawdata.csv**. *figshare*. Dataset. 2019.
<http://www.doi.org/10.6084/m9.figshare.8142992.v1>
10. Niederer M: **Validation with Reference samples 2017**. *figshare*. Dataset. 2019.
<http://www.doi.org/10.6084/m9.figshare.8143016.v1>
11. Niederer M: **Market survey 2019 rawdata**. *figshare*. Dataset. 2019.
<http://www.doi.org/10.6084/m9.figshare.8143031.v1>

Open Peer Review

Current Peer Review Status:

Version 1

Reviewer Report 03 July 2019

<https://doi.org/10.5256/f1000research.21163.r49712>

© 2019 Biedermann M. This is an open access peer review report distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Maurus Biedermann

Official Food Control Authority of the Canton of Zurich (Kantonales Labor), Zürich, Switzerland

The authors developed a rapid method for the detection of a nitrite treatment of tuna fish meat. It is a straight forward static headspace GC-MS method and may be most probably combined with the determination of carbon dioxide, another agent to stabilize the red color of the meat. Brining experiments with ¹⁵N labeled sodium nitrite proved the reaction path. The subject is of current importance and I suggest the indexing after minor revision.

Minor comments:

- Page 3, "introduction": Is there the possibility of combining the determination of CO and N₂O, e.g. even within the same GC run? If so, please mention that option/possibility.
- Page 3, "Samples and sample preparation": Please provide more details on the quantification, how was the calibration performed (one-point calibration, calibration curve, range of calibration...)?
- Page 4, "method development", first and second paragraph: Within the second paragraph of the chapter it was concluded N₂O to be the analyte in terms of robust chromatography. That explains the procedure of the two-step reduction described in the first paragraph. The GC temperature program was adjusted presumably based on the stronger retention of N₂O. Please revise order of chapters and/or conclusions.
- Page 5, figure 3 text: Please mention the value of the untreated samples (below LOQ?) within the figure text.
- Page 5, "Validation with reference samples": Please specify "free of N₂O"; according to the supplementary data the findings ranged between 13-31 ug/kg N₂O.

Is the rationale for developing the new method (or application) clearly explained?

Yes

Is the description of the method technically sound?

Yes

Are sufficient details provided to allow replication of the method development and its use by others?

Yes

If any results are presented, are all the source data underlying the results available to ensure full reproducibility?

Yes

Are the conclusions about the method and its performance adequately supported by the findings presented in the article?

Yes

Competing Interests: No competing interests were disclosed.

Reviewer Expertise: food and FCM analysis, GC-MS, on-line LC-GC, comprehensive two-dimensional GCxGC

I confirm that I have read this submission and believe that I have an appropriate level of expertise to confirm that it is of an acceptable scientific standard.

Author Response 07 Aug 2019

Markus Niederer, State Laboratory Basel-City, Kannenfeldstr. 2, Switzerland

We would like to thank the reviewer for his time and constructive feedback.

1. As requested, we now mention the possibility of combining the determination of CO and N₂O in the introduction.
2. We fulfilled the request by adding the expression „one-point calibration“.
3. Regarding the remarks on „method development“, we do not understand the line of thought: N₂O not NO is the target compound. Compared to CO, N₂O elutes only at higher oven temperatures. NO being a reactive intermediate is not amenable for GC analysis. With the first section explaining the fundamental differences between the detection of a CO treatment and a nitrite treatment and the second section providing details on the latter, we see no reason in revising the text.
4. The text of figure 3 was not changed because the required data is already mentioned in the text of the corresponding chapter.
5. “Validation with reference samples”: “free of N₂O” is specified as lower than 30 µg/kg.

Competing Interests: No competing interests were disclosed.

Reviewer Report 25 June 2019

<https://doi.org/10.5256/f1000research.21163.r49814>

© 2019 Busch U. This is an open access peer review report distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Ulrich Busch

The Bavarian State Office for Health and Food Safety (LGL), Erlangen, Germany

Addition of nitrite to tuna aiming at fresh reddish looking tuna has resulted in consumption of old tuna products with allergic reactions due to the formation of histamine. A systematic use of nitrite injections to rejuvenate tuna meat has been identified. The article "Identification of nitrite treated tuna fish meat via the determination of nitrous oxide by head space-gas chromatography/mass spectrometry" is well written, the method seems sufficiently valid to me. We do not know of an alternative method to prove nitrite treatment.

We would be very pleased if the manuscript could be indexed, as there is no published adequate method to prove the falsification.

Is the rationale for developing the new method (or application) clearly explained?

Yes

Is the description of the method technically sound?

Yes

Are sufficient details provided to allow replication of the method development and its use by others?

Yes

If any results are presented, are all the source data underlying the results available to ensure full reproducibility?

Yes

Are the conclusions about the method and its performance adequately supported by the findings presented in the article?

Yes

Competing Interests: No competing interests were disclosed.

Reviewer Expertise: food analytics molecular biology NMR, stable isotopes

I confirm that I have read this submission and believe that I have an appropriate level of expertise to confirm that it is of an acceptable scientific standard.

Author Response 07 Aug 2019

Markus Niederer, State Laboratory Basel-City, Kannenfeldstr. 2, Switzerland

We would like to thank the reviewer for his time and his important comments.

Competing Interests: No competing interests were disclosed.

The benefits of publishing with F1000Research:

- Your article is published within days, with no editorial bias
- You can publish traditional articles, null/negative results, case reports, data notes and more
- The peer review process is transparent and collaborative
- Your article is indexed in PubMed after passing peer review
- Dedicated customer support at every stage

For pre-submission enquiries, contact research@f1000.com

F1000Research